

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051122\
 Data File : P0085964.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 May 2022 00:31
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO051122AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 03:30:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 09:06:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.479	3.683	5355798	2299514	52.711	52.792
2)	SA Decachlor...	10.323	8.753	3173013	2047408	58.436	58.832
Target Compounds							
3)	L1 AR-1016-1	5.659	4.789	1405888	684815	439.906	438.489
4)	L1 AR-1016-2	5.682	4.809	1987964	911014	447.192	435.065
5)	L1 AR-1016-3	5.745	4.987	1263471	496444	460.796	449.121
6)	L1 AR-1016-4	5.844	5.030	1016441	398271	458.568	432.023
7)	L1 AR-1016-5	6.142	5.244	981730	522857	473.790	463.053
8)	L2 AR-1221-1	4.684	3.902	161325	69465	131.764	135.087
9)	L2 AR-1221-2	4.776	3.990	242171	99667	265.295	258.617
10)	L2 AR-1221-3	4.853	4.068	815135	356686	302.460	293.712
11)	L3 AR-1232-1	4.853	4.068	815135	356686	398.183	386.005
12)	L3 AR-1232-2	5.390	4.809	1027781	911014	1091.325	1117.588
13)	L3 AR-1232-3	5.682	4.987	1987964	496444	1131.013	1202.355
14)	L3 AR-1232-4	5.844	5.071	1016441	497855	1168.429	1328.704
15)	L3 AR-1232-5	5.936	5.244	846024	522857	1277.580	1248.422
16)	L4 AR-1242-1	5.659	4.789	1405888	684815	517.182	518.472
17)	L4 AR-1242-2	5.682	4.809	1987964	911014	528.807	519.519
18)	L4 AR-1242-3	5.745	4.987	1263471	496444	537.780	527.949
19)	L4 AR-1242-4	5.844	5.071	1016441	497855	536.247	527.821
20)	L4 AR-1242-5	6.586	5.598	1018523	620289	556.409	545.440
21)	L5 AR-1248-1	5.659	4.789	1405888	684815	696.033	705.263
22)	L5 AR-1248-2	5.936	5.030	846024	398271	313.596	299.449
23)	L5 AR-1248-3	6.142	5.071	981730	497855	330.877	361.771
24)	L5 AR-1248-4	6.548	5.244	978540	522857	303.979	326.549
25)	L5 AR-1248-5	6.586	5.640	1018523	525975	327.261	334.799
26)	L6 AR-1254-1	6.520	5.598	749413	620289	222.684	246.274
27)	L6 AR-1254-2	6.745	5.747	851484	147813	173.029	67.500 #
28)	L6 AR-1254-3	7.110	6.153	347987	205333	68.876	58.117
29)	L6 AR-1254-4	7.397	6.382	248462	150653	66.706	68.101
30)	L6 AR-1254-5	7.820	6.802	80230	47898	20.590	16.194
31)	L7 AR-1260-1	7.280	6.299	35466	71586	11.719	36.780 #
32)	L7 AR-1260-2	7.533	6.472	61124	22471	16.897	9.561 #
33)	L7 AR-1260-3	7.878	6.627	29620	22035	10.732	10.248
34)	L7 AR-1260-4	8.107	7.140f	42816	19911	12.690	11.058
35)	L7 AR-1260-5	0.000	7.319	0	8363	N.D.	1.931 #
36)	L8 AR-1262-1	7.878	6.802f	29620	47898	6.549	15.853 #
37)	L8 AR-1262-2	0.000	7.344	0	7077	N.D.	1.288 #
38)	L8 AR-1262-3	0.000	7.661	0	19152	N.D.	9.086 #
39)	L8 AR-1262-4	0.000	7.691	0	19128	N.D.	4.826 #
40)	L8 AR-1262-5	0.000	8.243f	0	17336	N.D.	9.677 #
41)	L9 AR-1268-1	0.000	7.661	0	19152	N.D.	3.139 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051122\
Data File : P0085964.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 May 2022 00:31
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 03:30:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	0.000	7.691	0	19128	N.D.	3.484 #
43) L9	AR-1268-3	0.000	7.922	0	20339	N.D.	4.475 #
44) L9	AR-1268-4	0.000	8.243f	0	17336	N.D.	8.879 #
45) L9	AR-1268-5	0.000	8.489	0	15792	N.D.	1.077 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051122\
Data File : P0085964.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 May 2022 00:31
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 03:30:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 5355798
Conc: 52.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.683 min
Delta R.T.: 0.000 min
Response: 2299514
Conc: 52.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.323 min
Delta R.T.: -0.002 min
Response: 3173013
Conc: 58.44 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.753 min
Delta R.T.: 0.000 min
Response: 2047408
Conc: 58.83 ng/ml



#3 AR-1016-1


R.T.: 5.659 min
Delta R.T.: 0.000 min
Response: 1405888
Conc: 439.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#3 AR-1016-1

R.T.: 4.789 min
Delta R.T.: 0.009 min
Response: 684815
Conc: 438.49 ng/ml




#4 AR-1016-2

R.T.: 5.682 min
Delta R.T.: 0.000 min
Response: 1987964
Conc: 447.19 ng/ml



#4 AR-1016-2


R.T.: 4.809 min
Delta R.T.: 0.009 min
Response: 911014
Conc: 435.06 ng/ml



#5 AR-1016-3


R.T.: 5.745 min
Delta R.T.: 0.000 min
Response: 1263471
Conc: 460.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#5 AR-1016-3

R.T.: 4.987 min
Delta R.T.: 0.002 min
Response: 496444
Conc: 449.12 ng/ml




#6 AR-1016-4

R.T.: 5.844 min
Delta R.T.: -0.001 min
Response: 1016441
Conc: 458.57 ng/ml



#6 AR-1016-4


R.T.: 5.030 min
Delta R.T.: 0.015 min
Response: 398271
Conc: 432.02 ng/ml



#7 AR-1016-5


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 981730
Conc: 473.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#7 AR-1016-5

R.T.: 5.244 min
Delta R.T.: 0.012 min
Response: 522857
Conc: 463.05 ng/ml




#8 AR-1221-1

R.T.: 4.684 min
Delta R.T.: -0.013 min
Response: 161325
Conc: 131.76 ng/ml



#8 AR-1221-1


R.T.: 3.902 min
Delta R.T.: -0.003 min
Response: 69465
Conc: 135.09 ng/ml



#9 AR-1221-2


R.T.: 4.776 min
Delta R.T.: -0.007 min
Response: 242171
Conc: 265.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#9 AR-1221-2

R.T.: 3.990 min
Delta R.T.: -0.002 min
Response: 99667
Conc: 258.62 ng/ml




#10 AR-1221-3

R.T.: 4.853 min
Delta R.T.: -0.008 min
Response: 815135
Conc: 302.46 ng/ml



#10 AR-1221-3


R.T.: 4.068 min
Delta R.T.: -0.002 min
Response: 356686
Conc: 293.71 ng/ml



#11 AR-1232-1


R.T.: 4.853 min
Delta R.T.: 0.000 min
Response: 815135
Conc: 398.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#11 AR-1232-1

R.T.: 4.068 min
Delta R.T.: 0.000 min
Response: 356686
Conc: 386.00 ng/ml




#12 AR-1232-2

R.T.: 5.390 min
Delta R.T.: 0.000 min
Response: 1027781
Conc: 1091.32 ng/ml



#12 AR-1232-2


R.T.: 4.809 min
Delta R.T.: -0.002 min
Response: 911014
Conc: 1117.59 ng/ml



#13 AR-1232-3


R.T.: 5.682 min
Delta R.T.: 0.000 min
Response: 1987964
Conc: 1131.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#13 AR-1232-3

R.T.: 4.987 min
Delta R.T.: -0.002 min
Response: 496444
Conc: 1202.36 ng/ml




#14 AR-1232-4

R.T.: 5.844 min
Delta R.T.: 0.000 min
Response: 1016441
Conc: 1168.43 ng/ml



#14 AR-1232-4


R.T.: 5.071 min
Delta R.T.: -0.002 min
Response: 497855
Conc: 1328.70 ng/ml



#15 AR-1232-5


R.T.: 5.936 min
Delta R.T.: 0.000 min
Response: 846024
Conc: 1277.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#15 AR-1232-5

R.T.: 5.244 min
Delta R.T.: -0.003 min
Response: 522857
Conc: 1248.42 ng/ml




#16 AR-1242-1

R.T.: 5.659 min
Delta R.T.: -0.001 min
Response: 1405888
Conc: 517.18 ng/ml



#16 AR-1242-1


R.T.: 4.789 min
Delta R.T.: -0.003 min
Response: 684815
Conc: 518.47 ng/ml



#17 AR-1242-2


R.T.: 5.682 min
Delta R.T.: -0.001 min
Response: 1987964
Conc: 528.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#17 AR-1242-2

R.T.: 4.809 min
Delta R.T.: -0.002 min
Response: 911014
Conc: 519.52 ng/ml




#18 AR-1242-3

R.T.: 5.745 min
Delta R.T.: 0.000 min
Response: 1263471
Conc: 537.78 ng/ml



#18 AR-1242-3


R.T.: 4.987 min
Delta R.T.: -0.002 min
Response: 496444
Conc: 527.95 ng/ml



#19 AR-1242-4


R.T.: 5.844 min
Delta R.T.: -0.002 min
Response: 1016441
Conc: 536.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#19 AR-1242-4

R.T.: 5.071 min
Delta R.T.: -0.003 min
Response: 497855
Conc: 527.82 ng/ml




#20 AR-1242-5

R.T.: 6.586 min
Delta R.T.: -0.001 min
Response: 1018523
Conc: 556.41 ng/ml



#20 AR-1242-5


R.T.: 5.598 min
Delta R.T.: -0.003 min
Response: 620289
Conc: 545.44 ng/ml



#21 AR-1248-1


R.T.: 5.659 min
Delta R.T.: 0.000 min
Response: 1405888
Conc: 696.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#21 AR-1248-1

R.T.: 4.789 min
Delta R.T.: -0.002 min
Response: 684815
Conc: 705.26 ng/ml




#22 AR-1248-2

R.T.: 5.936 min
Delta R.T.: -0.001 min
Response: 846024
Conc: 313.60 ng/ml



#22 AR-1248-2


R.T.: 5.030 min
Delta R.T.: -0.002 min
Response: 398271
Conc: 299.45 ng/ml



#23 AR-1248-3


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 981730
Conc: 330.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#23 AR-1248-3

R.T.: 5.071 min
Delta R.T.: -0.002 min
Response: 497855
Conc: 361.77 ng/ml




#24 AR-1248-4

R.T.: 6.548 min
Delta R.T.: 0.000 min
Response: 978540
Conc: 303.98 ng/ml



#24 AR-1248-4


R.T.: 5.244 min
Delta R.T.: -0.002 min
Response: 522857
Conc: 326.55 ng/ml



#25 AR-1248-5


R.T.: 6.586 min
Delta R.T.: -0.002 min
Response: 1018523
Conc: 327.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#25 AR-1248-5

R.T.: 5.640 min
Delta R.T.: -0.002 min
Response: 525975
Conc: 334.80 ng/ml




#26 AR-1254-1

R.T.: 6.520 min
Delta R.T.: -0.002 min
Response: 749413
Conc: 222.68 ng/ml



#26 AR-1254-1


R.T.: 5.598 min
Delta R.T.: -0.002 min
Response: 620289
Conc: 246.27 ng/ml



#27 AR-1254-2


R.T.: 6.745 min
Delta R.T.: 0.002 min
Response: 851484
Conc: 173.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#27 AR-1254-2

R.T.: 5.747 min
Delta R.T.: -0.001 min
Response: 147813
Conc: 67.50 ng/ml




#28 AR-1254-3

R.T.: 7.110 min
Delta R.T.: 0.000 min
Response: 347987
Conc: 68.88 ng/ml



#28 AR-1254-3


R.T.: 6.153 min
Delta R.T.: 0.000 min
Response: 205333
Conc: 58.12 ng/ml



#29 AR-1254-4


R.T.: 7.397 min
Delta R.T.: 0.000 min
Response: 248462
Conc: 66.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#29 AR-1254-4

R.T.: 6.382 min
Delta R.T.: -0.001 min
Response: 150653
Conc: 68.10 ng/ml




#30 AR-1254-5

R.T.: 7.820 min
Delta R.T.: 0.000 min
Response: 80230
Conc: 20.59 ng/ml



#30 AR-1254-5


R.T.: 6.802 min
Delta R.T.: -0.002 min
Response: 47898
Conc: 16.19 ng/ml



#31 AR-1260-1


R.T.: 7.280 min
Delta R.T.: 0.003 min
Response: 35466
Conc: 11.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#31 AR-1260-1

R.T.: 6.299 min
Delta R.T.: 0.011 min
Response: 71586
Conc: 36.78 ng/ml




#32 AR-1260-2

R.T.: 7.533 min
Delta R.T.: -0.002 min
Response: 61124
Conc: 16.90 ng/ml



#32 AR-1260-2


R.T.: 6.472 min
Delta R.T.: -0.004 min
Response: 22471
Conc: 9.56 ng/ml



#33 AR-1260-3


R.T.: 7.878 min
Delta R.T.: -0.019 min
Response: 29620
Conc: 10.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#33 AR-1260-3

R.T.: 6.627 min
Delta R.T.: -0.005 min
Response: 22035
Conc: 10.25 ng/ml




#34 AR-1260-4

R.T.: 8.107 min
Delta R.T.: -0.018 min
Response: 42816
Conc: 12.69 ng/ml



#34 AR-1260-4


R.T.: 7.140 min
Delta R.T.: 0.045 min
Response: 19911
Conc: 11.06 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 8.453 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#35 AR-1260-5

R.T.: 7.319 min
Delta R.T.: -0.011 min
Response: 8363
Conc: 1.93 ng/ml




#36 AR-1262-1

R.T.: 7.878 min
Delta R.T.: -0.017 min
Response: 29620
Conc: 6.55 ng/ml



#36 AR-1262-1


R.T.: 6.802 min
Delta R.T.: -0.040 min
Response: 47898
Conc: 15.85 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T.: 8.450 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#37 AR-1262-2

R.T.: 7.344 min
Delta R.T.: 0.000 min
Response: 7077
Conc: 1.29 ng/ml




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T.: 8.773 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.661 min
Delta R.T.: 0.032 min
Response: 19152
Conc: 9.09 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 8.864 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#39 AR-1262-4

R.T.: 7.691 min
Delta R.T.: -0.001 min
Response: 19128
Conc: 4.83 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T.: 9.539 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.243 min
Delta R.T.: 0.053 min
Response: 17336
Conc: 9.68 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T. : 8.770 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#41 AR-1268-1

R.T.: 7.661 min
Delta R.T.: 0.032 min
Response: 19152
Conc: 3.14 ng/ml




#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.867 min
Response: 0
Conc: N.D.



#42 AR-1268-2


R.T.: 7.691 min
Delta R.T.: -0.002 min
Response: 19128
Conc: 3.48 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 9.104 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242




#43 AR-1268-3

R.T.: 7.922 min
Delta R.T.: 0.019 min
Response: 20339
Conc: 4.47 ng/ml




#44 AR-1268-4

R.T.: 0.000 min
Exp R.T.: 9.541 min
Response: 0
Conc: N.D.



#44 AR-1268-4


R.T.: 8.243 min
Delta R.T.: 0.053 min
Response: 17336
Conc: 8.88 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.971 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
ICVPO051122AR1242



#45 AR-1268-5

R.T.: 8.489 min
Delta R.T.: -0.001 min
Response: 15792
Conc: 1.08 ng/ml