

Data Path : \\74.0.250.170\terastorage\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051219\
 Data File : P0056156.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 May 2019 15:40
 Operator : SM/SJ
 Sample : K2767-02
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 P026-SS010-0002-01

Manual Integrations
 APPROVED

Ankita
 5/15/2019 12:27:45 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 12 21:50:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 03 15:20:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.263	3.580	618301	580786	20.661	20.564
2) SA Decachlor...	9.851	8.523	579544	426964	15.489	15.788m
Target Compounds						
31) L7 AR-1260-1	7.015	6.117	641927	549493	297.258	249.408
32) L7 AR-1260-2	7.271	6.303	888960	823645	339.442	269.732m
33) L7 AR-1260-3	7.627	6.456	644817	493531	322.410	198.044m#
34) L7 AR-1260-4	7.851	6.924	654759	437514	283.126	212.028 #
35) L7 AR-1260-5	8.161	7.165	933780	1119219	220.111	222.851

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : \\74.0.250.170\terastorage\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051219\
Data File : PO056156.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 May 2019 15:40
Operator : SM/SJ
Sample : K2767-02
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
P026-SS010-0002-01

Manual Integrations
APPROVED

Ankita
5/15/2019 12:27:45 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 12 21:50:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 03 15:20:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

