

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051222\
 Data File : P0085986.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 May 2022 20:36
 Operator : YP\AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 07:09:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 12 05:35:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.479	3.682	4808536	2033817	47.351	46.665
2)	SA Decachlor...	10.329	8.747	2766221	1596279	52.200	47.031
Target Compounds							
16)	L4 AR-1242-1	5.660	4.787	1341702	591865	493.570	448.100
17)	L4 AR-1242-2	5.683	4.807	1863914	781264	495.809	445.527
18)	L4 AR-1242-3	5.746	4.984	1169059	424488	497.595	451.427
19)	L4 AR-1242-4	5.846	5.068	947202	425599	499.719	451.215
20)	L4 AR-1242-5	6.589	5.595	888552	498983	485.407	438.771



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051222\
Data File : P0085986.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 May 2022 20:36
Operator : YP\AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 07:09:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 12 05:35:26 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 4808536
Conc: 47.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: -0.002 min
Response: 2033817
Conc: 46.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.329 min
Delta R.T.: 0.002 min
Response: 2766221
Conc: 52.20 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.747 min
Delta R.T.: -0.012 min
Response: 1596279
Conc: 47.03 ng/ml



#16 AR-1242-1


R.T.: 5.660 min
Delta R.T.: 0.000 min
Response: 1341702
Conc: 493.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.787 min
Delta R.T.: -0.005 min
Response: 591865
Conc: 448.10 ng/ml




#17 AR-1242-2

R.T.: 5.683 min
Delta R.T.: 0.000 min
Response: 1863914
Conc: 495.81 ng/ml



#17 AR-1242-2


R.T.: 4.807 min
Delta R.T.: -0.004 min
Response: 781264
Conc: 445.53 ng/ml



#18 AR-1242-3


R.T.: 5.746 min
Delta R.T.: 0.000 min
Response: 1169059
Conc: 497.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.984 min
Delta R.T.: -0.006 min
Response: 424488
Conc: 451.43 ng/ml




#19 AR-1242-4

R.T.: 5.846 min
Delta R.T.: 0.000 min
Response: 947202
Conc: 499.72 ng/ml



#19 AR-1242-4


R.T.: 5.068 min
Delta R.T.: -0.006 min
Response: 425599
Conc: 451.22 ng/ml



#20 AR-1242-5

R.T.: 6.589 min
Delta R.T.: 0.001 min
Response: 888552
Conc: 485.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.595 min
Delta R.T.: -0.006 min
Response: 498983
Conc: 438.77 ng/ml