

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051222\
 Data File : P0086003.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 May 2022 1:44
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 07:16:01 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 12 05:35:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.478	3.679	6090140	2492319	59.971	57.186
2) SA Decachlor...	10.332	8.746	2652976	1471343	50.063	43.350
Target Compounds						
26) L6 AR-1254-1	6.524	5.595	1830641	1176098	543.965	466.947
27) L6 AR-1254-2	6.745	5.743	2585627	997455	525.423	455.494
28) L6 AR-1254-3	7.112	6.149	2724574	1543263	539.263	436.803
29) L6 AR-1254-4	7.401	6.378	1922881	959080	516.249	433.539
30) L6 AR-1254-5	7.823	6.798	1811132	1216867	464.812	411.404



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051222\
Data File : P0086003.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 May 2022 1:44
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 07:16:01 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 12 05:35:26 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.478 min
Delta R.T.: 0.000 min
Response: 6090140
Conc: 59.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.679 min
Delta R.T.: -0.005 min
Response: 2492319
Conc: 57.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.332 min
Delta R.T.: 0.005 min
Response: 2652976
Conc: 50.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.746 min
Delta R.T.: -0.013 min
Response: 1471343
Conc: 43.35 ng/ml



#26 AR-1254-1


R.T.: 6.524 min
Delta R.T.: 0.002 min
Response: 1830641
Conc: 543.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.595 min
Delta R.T.: -0.005 min
Response: 1176098
Conc: 466.95 ng/ml




#27 AR-1254-2

R.T.: 6.745 min
Delta R.T.: 0.002 min
Response: 2585627
Conc: 525.42 ng/ml



#27 AR-1254-2


R.T.: 5.743 min
Delta R.T.: -0.005 min
Response: 997455
Conc: 455.49 ng/ml



#28 AR-1254-3


R.T.: 7.112 min
Delta R.T.: 0.001 min
Response: 2724574
Conc: 539.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.149 min
Delta R.T.: -0.005 min
Response: 1543263
Conc: 436.80 ng/ml




#29 AR-1254-4

R.T.: 7.401 min
Delta R.T.: 0.003 min
Response: 1922881
Conc: 516.25 ng/ml



#29 AR-1254-4


R.T.: 6.378 min
Delta R.T.: -0.005 min
Response: 959080
Conc: 433.54 ng/ml



#30 AR-1254-5

R.T.: 7.823 min
Delta R.T.: 0.003 min
Response: 1811132
Conc: 464.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.798 min
Delta R.T.: -0.006 min
Response: 1216867
Conc: 411.40 ng/ml