

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051225\
 Data File : P0111007.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 May 2025 09:44
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 01:49:59 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 02 23:47:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.683	3.679	320.4E6	284.8E6	56.005	58.802
2) SA Decachlor...	8.722	8.671	261.5E6	93280438	42.853	55.285 #
Target Compounds						
21) L5 AR-1248-1	4.773	4.757	69515565	58381684	529.190	605.776
22) L5 AR-1248-2	5.011	4.994	93152551	80892815	531.594	598.666
23) L5 AR-1248-3	5.225	5.035	115.0E6	85284692	521.655	587.591
24) L5 AR-1248-4	5.580	5.206	162.8E6	99401559	496.892	593.385
25) L5 AR-1248-5	5.621	5.598	117.3E6	98422226	502.586	586.270



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051225\
Data File : PO111007.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 May 2025 09:44
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 01:49:59 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050225.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 02 23:47:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.683 min
Delta R.T.: 0.000 min
Response: 320394065
Conc: 56.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.679 min
Delta R.T.: -0.001 min
Response: 284783718
Conc: 58.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.722 min
Delta R.T.: -0.003 min
Response: 261509108
Conc: 42.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.671 min
Delta R.T.: -0.005 min
Response: 93280438
Conc: 55.29 ng/ml



#21 AR-1248-1


R.T.: 4.773 min
Delta R.T.: -0.002 min
Response: 69515565
Conc: 529.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.757 min
Delta R.T.: -0.003 min
Response: 58381684
Conc: 605.78 ng/ml




#22 AR-1248-2

R.T.: 5.011 min
Delta R.T.: -0.003 min
Response: 93152551
Conc: 531.59 ng/ml



#22 AR-1248-2


R.T.: 4.994 min
Delta R.T.: -0.003 min
Response: 80892815
Conc: 598.67 ng/ml



#23 AR-1248-3


R.T.: 5.225 min
Delta R.T.: -0.003 min
Response: 115008087
Conc: 521.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.035 min
Delta R.T.: -0.004 min
Response: 85284692
Conc: 587.59 ng/ml




#24 AR-1248-4

R.T.: 5.580 min
Delta R.T.: -0.003 min
Response: 162838512
Conc: 496.89 ng/ml



#24 AR-1248-4


R.T.: 5.206 min
Delta R.T.: -0.003 min
Response: 99401559
Conc: 593.38 ng/ml



#25 AR-1248-5

R.T.: 5.621 min
Delta R.T.: -0.002 min
Response: 117327581
Conc: 502.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.598 min
Delta R.T.: -0.003 min
Response: 98422226
Conc: 586.27 ng/ml