

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051321\
 Data File : P0077749.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 May 2021 14:21
 Operator : DD\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 15:21:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 13 15:20:38 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.355	3.633	3024998	1308994	50.000	50.000
2)	SA Decachlor...	9.979	8.832	2059030	1182016	50.000	50.000
Target Compounds							
3)	L1 AR-1016-1	5.652	4.866	979601	469736	500.000	500.000
4)	L1 AR-1016-2	5.674	4.884	1450227	676693	500.000	500.000
5)	L1 AR-1016-3	5.738	5.072	878640	355496	500.000	500.000
6)	L1 AR-1016-4	5.843	5.125	703857	298533	500.000	500.000
7)	L1 AR-1016-5	6.153	5.349	693034	371113	500.000	500.000
31)	L7 AR-1260-1	7.315	6.430	1260504	713848	500.000	500.000
32)	L7 AR-1260-2	7.579	6.627	1414860	860470	500.000	500.000
33)	L7 AR-1260-3	7.940	6.778	945612	796172	500.000	500.000
34)	L7 AR-1260-4	8.166	7.256	1115771	594309	500.000	500.000
35)	L7 AR-1260-5	8.479	7.503	2047565	1396297	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051321\
Data File : P0077749.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 May 2021 14:21
Operator : DD\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 15:21:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051321.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 13 15:20:38 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.355 min
Delta R.T.: 0.000 min
Response: 3024998
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.633 min
Delta R.T.: 0.000 min
Response: 1308994
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.979 min
Delta R.T.: 0.000 min
Response: 2059030
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.832 min
Delta R.T.: 0.000 min
Response: 1182016
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.652 min
Delta R.T.: 0.000 min
Response: 979601
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.866 min
Delta R.T.: 0.000 min
Response: 469736
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.674 min
Delta R.T.: 0.000 min
Response: 1450227
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.884 min
Delta R.T.: 0.000 min
Response: 676693
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.738 min
Delta R.T.: 0.000 min
Response: 878640
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 5.072 min
Delta R.T.: 0.000 min
Response: 355496
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.843 min
Delta R.T.: 0.000 min
Response: 703857
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 5.125 min
Delta R.T.: 0.000 min
Response: 298533
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.153 min
Delta R.T.: 0.000 min
Response: 693034
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 371113
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.315 min
Delta R.T.: 0.000 min
Response: 1260504
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.430 min
Delta R.T.: 0.000 min
Response: 713848
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.579 min
Delta R.T.: 0.000 min
Response: 1414860
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.627 min
Delta R.T.: 0.000 min
Response: 860470
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.940 min
Delta R.T.: 0.000 min
Response: 945612
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.778 min
Delta R.T.: 0.000 min
Response: 796172
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.166 min
Delta R.T.: 0.000 min
Response: 1115771
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.256 min
Delta R.T.: 0.000 min
Response: 594309
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.479 min
Delta R.T.: 0.000 min
Response: 2047565
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.503 min
Delta R.T.: 0.000 min
Response: 1396297
Conc: 500.00 ng/ml