

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051325\
 Data File : P0111030.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 10:28
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 23:19:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 02 23:47:49 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.684	3.680	92205092	77249742	16.118	15.950
2) SA Decachlor...	8.723	8.672	80433913	27530220	13.181	16.317

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051325\
Data File : PO111030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 10:28
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 23:19:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050225.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 02 23:47:49 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 92205092
Conc: 16.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.680 min
Delta R.T.: 0.000 min
Response: 77249742
Conc: 15.95 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.723 min
Delta R.T.: -0.002 min
Response: 80433913
Conc: 13.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.672 min
Delta R.T.: -0.003 min
Response: 27530220
Conc: 16.32 ng/ml