

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0051421\
 Data File : P0077814.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 May 2021 14:34
 Operator : DD\AJ
 Sample : M2262-03
 Misc : AR1232 MDL 25 PPB
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 MDL-SOIL-03-QT2-2021

Manual Integrations
 APPROVED

Ankita
 5/18/2021 8:55:39 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 14 15:43:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0051321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 14 05:52:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.352	3.631	1196043	510289	20.375	20.054
2) SA Decachlor...	9.971	8.828	852149	536682	20.613	22.331
Target Compounds						
11) L3 AR-1232-1	4.784	4.067	44058	15043	33.405	26.273
12) L3 AR-1232-2	5.355	4.882	40685	21869	61.028	37.858 #
13) L3 AR-1232-3	5.672	5.071	28685	7191	23.833	25.166
14) L3 AR-1232-4	5.838	5.164	13680	9321	24.717m	34.954 #
15) L3 AR-1232-5	5.943	5.347	12599	8498	31.722m	30.277

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051421\
Data File : P0077814.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 May 2021 14:34
Operator : DD\AJ
Sample : M2262-03
Misc : AR1232 MDL 25 PPB
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MDL-SOIL-03-QT2-2021

Manual Integrations
APPROVED

Ankita
5/18/2021 8:55:39 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 14 15:43:50 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 14 05:52:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

