

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0051421\
 Data File : P0077822.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 May 2021 16:52
 Operator : DD\AJ
 Sample : M2262-09
 Misc : AR1242 MDL 25 PPB
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MDL-WATER-03-QT2-2021

Manual Integrations
 APPROVED

Ankita
 5/18/2021 8:55:44 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 14 17:31:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0051321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 14 05:52:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.351	3.630	1305296	610605	22.237	23.997m
2)	SA Decachlor...	9.972	8.827	965386	606433	23.352	25.234
Target Compounds							
16)	L4 AR-1242-1	5.647	4.861	30329	26041	20.236	35.284 #
17)	L4 AR-1242-2	5.670	4.881	61212	36129	27.125	34.621 #
18)	L4 AR-1242-3	5.733	5.069	46759	14151	33.764	26.132
19)	L4 AR-1242-4	5.839	5.164	25825	22657	23.908	40.032 #
20)	L4 AR-1242-5	6.610	5.720	30546	17107	26.845	24.876

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051421\
Data File : P0077822.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 May 2021 16:52
Operator : DD\AJ
Sample : M2262-09
Misc : AR1242 MDL 25 PPB
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MDL-WATER-03-QT2-2021

Manual Integrations
APPROVED

Ankita
5/18/2021 8:55:44 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 14 17:31:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 14 05:52:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

