

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051421\
 Data File : P0077825.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 May 2021 17:43
 Operator : DD\AJ
 Sample : M2262-09
 Misc : AR1262 MDL 25 PPB
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MDL-WATER-03-QT2-2021

Manual Integrations
 APPROVED

Ankita
 5/18/2021 8:55:46 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 14 18:37:54 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 14 05:52:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.351	3.631	1383565	605700	23.570	23.804
2) SA Decachlor...	9.971	8.826	1005345	629928	24.318	26.211
Target Compounds						
36) L8 AR-1262-1	7.933	6.951	84802	31560	26.997m	31.918
37) L8 AR-1262-2	8.473	7.498	135106	102210	26.288	30.695
38) L8 AR-1262-3	8.746	7.782	96635	40658	28.089m	29.460
39) L8 AR-1262-4	8.823	7.844	93324	77605	59.476	31.106 #
40) L8 AR-1262-5	9.374	8.339	49817	32427	28.423	27.938

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051421\
Data File : PO077825.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 May 2021 17:43
Operator : DD\AJ
Sample : M2262-09
Misc : AR1262 MDL 25 PPB
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
MDL-WATER-03-QT2-2021

Manual Integrations
APPROVED

Ankita
5/18/2021 8:55:46 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 14 18:37:54 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 14 05:52:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

