

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051523\
 Data File : P0095047.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2023 14:49
 Operator : YP/AJ
 Sample : 02784-01MS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BU-03-051223MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/16/2023
 Supervised By :Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 17:30:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.414	3.614	68930192	29077303	20.011	21.731m
2)	SA Decachlor...	10.252	8.623	58766036	25032880	24.215	22.240
Target Compounds							
3)	L1 AR-1016-1	5.592	4.696	50033711	16517360	491.162	416.457
4)	L1 AR-1016-2	5.614	4.713	67680275	25227139	447.307	426.738
5)	L1 AR-1016-3	5.677	4.888	40448424	17215288	407.975	548.402 #
6)	L1 AR-1016-4	5.775	4.933	32998271	14046335	443.338	485.757
7)	L1 AR-1016-5	6.073	5.144	42493620	16611229	509.316	453.878
31)	L7 AR-1260-1	7.212	6.178	68259113	32612868	460.373	463.249
32)	L7 AR-1260-2	7.470	6.367	122.5E6	36785192	755.462	460.140 #
33)	L7 AR-1260-3	7.837	6.519	50724860	34097164	388.002	453.403
34)	L7 AR-1260-4	8.064	6.992	62122141	25713818	444.261	410.768
35)	L7 AR-1260-5	8.393	7.237	115.5E6	57685765	471.911	457.683

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051523\
Data File : P0095047.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2023 14:49
Operator : YP/AJ
Sample : 02784-01MS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BU-03-051223MS

Manual Integrations
APPROVED

Reviewed By : Yogesh Patel 05/16/2023
Supervised By : Ankita Jodhani 05/16/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 17:30:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

