

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051524\
 Data File : P0103791.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2024 19:04
 Operator : YP/AJ
 Sample : P2403-07
 Misc : AR1254 LOD 25 PPB
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOD-MDL-WATER-01-QT2-2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 16 01:49:21 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 07 04:55:35 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.450	3.584	129.1E6	44088936	16.772	17.501
2)	SA Decachlor...	10.208	8.540	86967572	21954984	18.347	16.877
Target Compounds							
26)	L6 AR-1254-1	6.474	5.454	6103756	2814564	24.786	23.092
27)	L6 AR-1254-2	6.691	5.600	7857232	2640056	21.708	25.000
28)	L6 AR-1254-3	7.057	5.999	8102353	3728020	22.256	21.691
29)	L6 AR-1254-4	7.340	6.225	6199344	2192179	21.871	21.272
30)	L6 AR-1254-5	7.759	6.640	5419395	2847369	20.336	20.636

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051524\
Data File : PO103791.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2024 19:04
Operator : YP/AJ
Sample : P2403-07
Misc : AR1254 LOD 25 PPB
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-01-QT2-2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 16 01:49:21 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 07 04:55:35 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

