

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051525\
 Data File : P0111088.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 May 2025 03:03
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO051525AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 06:40:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 15 06:36:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.683	3.680	302.3E6	269.4E6	51.383	49.416
2) SA Decachlor...	8.725	8.674	257.8E6	92861317	49.936	50.191
Target Compounds						
21) L5 AR-1248-1	4.773	4.758	75013440	61783121	489.753	476.251
22) L5 AR-1248-2	5.012	4.994	98705358	84886979	492.253	482.609
23) L5 AR-1248-3	5.226	5.037	123.6E6	90239307	482.709	480.983
24) L5 AR-1248-4	5.580	5.207	175.1E6	105.4E6	483.415	479.447
25) L5 AR-1248-5	5.622	5.598	126.5E6	105.1E6	482.660	481.511



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051525\
Data File : PO111088.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 May 2025 03:03
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO051525AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 06:40:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051525.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 15 06:36:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.683 min
Delta R.T.: 0.000 min
Response: 302294556
Conc: 51.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051525AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.680 min
Delta R.T.: 0.000 min
Response: 269387714
Conc: 49.42 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.725 min
Delta R.T.: 0.000 min
Response: 257811098
Conc: 49.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.674 min
Delta R.T.: 0.000 min
Response: 92861317
Conc: 50.19 ng/ml



#21 AR-1248-1


R.T.: 4.773 min
Delta R.T.: 0.000 min
Response: 75013440
Conc: 489.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051525AR1248




#21 AR-1248-1

R.T.: 4.758 min
Delta R.T.: 0.000 min
Response: 61783121
Conc: 476.25 ng/ml




#22 AR-1248-2

R.T.: 5.012 min
Delta R.T.: 0.000 min
Response: 98705358
Conc: 492.25 ng/ml



#22 AR-1248-2


R.T.: 4.994 min
Delta R.T.: 0.000 min
Response: 84886979
Conc: 482.61 ng/ml



#23 AR-1248-3


R.T.: 5.226 min
Delta R.T.: 0.000 min
Response: 123575911
Conc: 482.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051525AR1248




#23 AR-1248-3

R.T.: 5.037 min
Delta R.T.: 0.000 min
Response: 90239307
Conc: 480.98 ng/ml




#24 AR-1248-4

R.T.: 5.580 min
Delta R.T.: 0.000 min
Response: 175083679
Conc: 483.41 ng/ml



#24 AR-1248-4


R.T.: 5.207 min
Delta R.T.: 0.000 min
Response: 105401554
Conc: 479.45 ng/ml



#25 AR-1248-5

R.T.: 5.622 min
Delta R.T.: 0.000 min
Response: 126483558
Conc: 482.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO051525AR1248



#25 AR-1248-5

R.T.: 5.598 min
Delta R.T.: 0.000 min
Response: 105125933
Conc: 481.51 ng/ml