

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051619\
 Data File : PD053280.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 May 2019 17:10
 Operator : SM\AJ
 Sample : TOXAPH401
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 17 10:49:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051619CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri May 17 10:42:14 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.458	3.991	27715620	33336811	38.986	39.084
27)	SA Decachlor...	8.234	9.056	44703058	51843390	79.897	76.636
Target Compounds							
22)	Toxaphene-1	5.723	6.377	18202652	14688210	3944.002	3777.202
23)	Toxaphene-2	5.884	6.667	16783015	18371184	3924.603	3858.075
24)	Toxaphene-3	6.209	7.073	17431875	24078123	4073.740	3820.891
25)	Toxaphene-4	6.330	7.149	39896791	69712202	3830.731	3847.219
26)	Toxaphene-5	7.066	7.648	38891192	31274057	3741.396	3864.531

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_D\Data\PD051619\
Data File : PD053280.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 May 2019 17:10
Operator : SM\AJ
Sample : TOXAPH401
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 17 10:49:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_D\Method\PD051619CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 10:42:14 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Instrument :
ECD_O
ClientSampleId :