

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0051619\
 Data File : P0056286.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 May 2019 12:46
 Operator : SM/SJ
 Sample : K2836-02MS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CULVER-CBTC-WC1-051019MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 16 13:01:47 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0050319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 03 15:20:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.266	3.577	916600	850811	30.629	30.124
2)	SA Decachlor...	9.854	8.523	822022	625377	21.969	23.125
Target Compounds							
3)	L1 AR-1016-1	5.449	4.665	359816	277470	253.192	206.464
4)	L1 AR-1016-2	5.449	4.665	359816	277470	181.829	151.959
5)	L1 AR-1016-3	5.510	4.839	217951	159698	170.043	159.359
6)	L1 AR-1016-4	5.609	4.921	178752	130517	171.018	160.412
7)	L1 AR-1016-5	5.900	5.092	179005	156698	158.901	142.342
9)	L2 AR-1221-2	4.565	0.000	37015	0	110.381	N.D. #
10)	L2 AR-1221-3	4.639	0.000	126315	0	125.613	N.D. #
11)	L3 AR-1232-1	4.639	0.000	126315	0	147.065	N.D. #
12)	L3 AR-1232-2	5.161	4.665	174201	277470	359.092	316.584
13)	L3 AR-1232-3	5.449	4.839	359816	159698	371.908	332.934
14)	L3 AR-1232-4	5.609	4.921	178752	130517	351.501	307.269
15)	L3 AR-1232-5	5.698	5.092	168228	156698	387.466	323.351
16)	L4 AR-1242-1	5.449	4.665	359816	277470	297.173	247.603
17)	L4 AR-1242-2	5.449	4.665	359816	277470	213.111	182.372
18)	L4 AR-1242-3	5.510	4.839	217951	159698	198.944	184.406
19)	L4 AR-1242-4	5.609	4.921	178752	130517	198.653	155.803
20)	L4 AR-1242-5	6.340	0.000	29612	0	27.068	N.D. #
21)	L5 AR-1248-1	5.449	4.665	359816	277470	397.534	330.901
22)	L5 AR-1248-2	5.698	4.921	168228	130517	126.144	118.785
23)	L5 AR-1248-3	5.900	4.921	179005	130517	121.217	114.978
24)	L5 AR-1248-4	6.303	5.092	34944	156698	20.040	111.594 #
25)	L5 AR-1248-5	6.340	0.000	29612	0	17.728	N.D. #
26)	L6 AR-1254-1	6.274	0.000	144704	0	68.912	N.D. #
27)	L6 AR-1254-2	6.490	0.000	160226	0	48.278	N.D. #
28)	L6 AR-1254-3	6.857	6.000	126344	214754	34.753	54.127 #
29)	L6 AR-1254-4	7.139	0.000	59538	0	21.883	N.D. #
30)	L6 AR-1254-5	7.556	6.628	474222	419350	165.458	115.885 #
31)	L7 AR-1260-1	7.018	6.114	353502	315771	163.697	143.324
32)	L7 AR-1260-2	7.273	6.302	413014	406262	157.706	133.045
33)	L7 AR-1260-3	7.630	6.454	258944	349613	129.473	140.293
34)	L7 AR-1260-4	7.855	6.922	315842	238711	136.574	115.684
35)	L7 AR-1260-5	8.164	7.164	525109	554594	123.779	110.427
36)	L8 AR-1262-1	7.630	6.663	258944	278074	80.772	67.810
37)	L8 AR-1262-2	8.164	7.164	525109	554594	97.809	88.695
38)	L8 AR-1262-3	8.474	7.508	388628	387051	107.094	157.148 #
39)	L8 AR-1262-4	8.523	7.508	208133	387051	76.658	94.519
40)	L8 AR-1262-5	9.162	0.000	119747	0	66.286	N.D. #
41)	L9 AR-1268-1	8.474	7.508	388628	387051	64.355	55.862
42)	L9 AR-1268-2	8.523	7.508	208133	387051	38.024	68.033 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051619\
Data File : P0056286.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 May 2019 12:46
Operator : SM/SJ
Sample : K2836-02MS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 16 13:01:47 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 03 15:20:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
44)	L9 AR-1268-4	9.162	0.000	119747	0	61.027	N.D. #
45)	L9 AR-1268-5	9.544	0.000	46819	0	3.470	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051619\
Data File : P0056286.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 May 2019 12:46
Operator : SM/SJ
Sample : K2836-02MS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 16 13:01:47 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 03 15:20:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.266 min
Delta R.T.: -0.013 min
Response: 916600
Conc: 30.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#1 Tetrachloro-m-xylene

R.T.: 3.577 min
Delta R.T.: -0.016 min
Response: 850811
Conc: 30.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.854 min
Delta R.T.: -0.025 min
Response: 822022
Conc: 21.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.523 min
Delta R.T.: -0.020 min
Response: 625377
Conc: 23.13 ng/ml



#3 AR-1016-1


R.T.: 5.449 min
Delta R.T.: 0.009 min
Response: 359816
Conc: 253.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#3 AR-1016-1

R.T.: 4.665 min
Delta R.T.: 0.000 min
Response: 277470
Conc: 206.46 ng/ml




#4 AR-1016-2

R.T.: 5.449 min
Delta R.T.: -0.013 min
Response: 359816
Conc: 181.83 ng/ml



#4 AR-1016-2


R.T.: 4.665 min
Delta R.T.: -0.018 min
Response: 277470
Conc: 151.96 ng/ml



#5 AR-1016-3


R.T.: 5.510 min
Delta R.T.: -0.013 min
Response: 217951
Conc: 170.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#5 AR-1016-3

R.T.: 4.839 min
Delta R.T.: -0.018 min
Response: 159698
Conc: 159.36 ng/ml




#6 AR-1016-4

R.T.: 5.609 min
Delta R.T.: -0.013 min
Response: 178752
Conc: 171.02 ng/ml



#6 AR-1016-4


R.T.: 4.921 min
Delta R.T.: 0.022 min
Response: 130517
Conc: 160.41 ng/ml



#7 AR-1016-5


R.T.: 5.900 min
Delta R.T.: -0.013 min
Response: 179005
Conc: 158.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#7 AR-1016-5

R.T.: 5.092 min
Delta R.T.: -0.017 min
Response: 156698
Conc: 142.34 ng/ml




#9 AR-1221-2

R.T.: 4.565 min
Delta R.T.: -0.013 min
Response: 37015
Conc: 110.38 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 3.888 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 4.639 min
Delta R.T.: -0.015 min
Response: 126315
Conc: 125.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 3.964 min
Response: 0
Conc: N.D.




#11 AR-1232-1

R.T.: 4.639 min
Delta R.T.: -0.011 min
Response: 126315
Conc: 147.07 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T. : 3.963 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 5.161 min
Delta R.T.: -0.013 min
Response: 174201
Conc: 359.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#12 AR-1232-2

R.T.: 4.665 min
Delta R.T.: -0.018 min
Response: 277470
Conc: 316.58 ng/ml




#13 AR-1232-3

R.T.: 5.449 min
Delta R.T.: -0.013 min
Response: 359816
Conc: 371.91 ng/ml



#13 AR-1232-3


R.T.: 4.839 min
Delta R.T.: -0.018 min
Response: 159698
Conc: 332.93 ng/ml



#14 AR-1232-4


R.T.: 5.609 min
Delta R.T.: -0.013 min
Response: 178752
Conc: 351.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#14 AR-1232-4

R.T.: 4.921 min
Delta R.T.: -0.018 min
Response: 130517
Conc: 307.27 ng/ml




#15 AR-1232-5

R.T.: 5.698 min
Delta R.T.: -0.013 min
Response: 168228
Conc: 387.47 ng/ml



#15 AR-1232-5


R.T.: 5.092 min
Delta R.T.: -0.018 min
Response: 156698
Conc: 323.35 ng/ml



#16 AR-1242-1


R.T.: 5.449 min
Delta R.T.: 0.010 min
Response: 359816
Conc: 297.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#16 AR-1242-1

R.T.: 4.665 min
Delta R.T.: 0.001 min
Response: 277470
Conc: 247.60 ng/ml




#17 AR-1242-2

R.T.: 5.449 min
Delta R.T.: -0.012 min
Response: 359816
Conc: 213.11 ng/ml



#17 AR-1242-2


R.T.: 4.665 min
Delta R.T.: -0.017 min
Response: 277470
Conc: 182.37 ng/ml



#18 AR-1242-3


R.T.: 5.510 min
Delta R.T.: -0.012 min
Response: 217951
Conc: 198.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#18 AR-1242-3

R.T.: 4.839 min
Delta R.T.: -0.018 min
Response: 159698
Conc: 184.41 ng/ml




#19 AR-1242-4

R.T.: 5.609 min
Delta R.T.: -0.012 min
Response: 178752
Conc: 198.65 ng/ml



#19 AR-1242-4


R.T.: 4.921 min
Delta R.T.: -0.017 min
Response: 130517
Conc: 155.80 ng/ml



#20 AR-1242-5


R.T.: 6.340 min
Delta R.T.: -0.012 min
Response: 29612
Conc: 27.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 5.457 min
Response: 0
Conc: N.D.




#21 AR-1248-1

R.T.: 5.449 min
Delta R.T.: 0.010 min
Response: 359816
Conc: 397.53 ng/ml



#21 AR-1248-1


R.T.: 4.665 min
Delta R.T.: 0.001 min
Response: 277470
Conc: 330.90 ng/ml



#22 AR-1248-2


R.T.: 5.698 min
Delta R.T.: -0.012 min
Response: 168228
Conc: 126.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#22 AR-1248-2

R.T.: 4.921 min
Delta R.T.: 0.024 min
Response: 130517
Conc: 118.79 ng/ml




#23 AR-1248-3

R.T.: 5.900 min
Delta R.T.: -0.012 min
Response: 179005
Conc: 121.22 ng/ml



#23 AR-1248-3


R.T.: 4.921 min
Delta R.T.: -0.017 min
Response: 130517
Conc: 114.98 ng/ml



#24 AR-1248-4


R.T.: 6.303 min
Delta R.T.: -0.011 min
Response: 34944
Conc: 20.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#24 AR-1248-4

R.T.: 5.092 min
Delta R.T.: -0.017 min
Response: 156698
Conc: 111.59 ng/ml




#25 AR-1248-5

R.T.: 6.340 min
Delta R.T.: -0.011 min
Response: 29612
Conc: 17.73 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 5.498 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 6.274 min
Delta R.T.: -0.013 min
Response: 144704
Conc: 68.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 5.457 min
Response: 0
Conc: N.D.




#27 AR-1254-2

R.T.: 6.490 min
Delta R.T.: -0.014 min
Response: 160226
Conc: 48.28 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T. : 5.603 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 6.857 min
Delta R.T.: -0.012 min
Response: 126344
Conc: 34.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#28 AR-1254-3

R.T.: 6.000 min
Delta R.T.: -0.003 min
Response: 214754
Conc: 54.13 ng/ml




#29 AR-1254-4

R.T.: 7.139 min
Delta R.T.: -0.015 min
Response: 59538
Conc: 21.88 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 6.231 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 7.556 min
Delta R.T.: -0.014 min
Response: 474222
Conc: 165.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#30 AR-1254-5

R.T.: 6.628 min
Delta R.T.: -0.018 min
Response: 419350
Conc: 115.88 ng/ml




#31 AR-1260-1

R.T.: 7.018 min
Delta R.T.: -0.015 min
Response: 353502
Conc: 163.70 ng/ml



#31 AR-1260-1


R.T.: 6.114 min
Delta R.T.: -0.020 min
Response: 315771
Conc: 143.32 ng/ml



#32 AR-1260-2


R.T.: 7.273 min
Delta R.T.: -0.015 min
Response: 413014
Conc: 157.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#32 AR-1260-2

R.T.: 6.302 min
Delta R.T.: -0.019 min
Response: 406262
Conc: 133.04 ng/ml




#33 AR-1260-3

R.T.: 7.630 min
Delta R.T.: -0.015 min
Response: 258944
Conc: 129.47 ng/ml



#33 AR-1260-3


R.T.: 6.454 min
Delta R.T.: -0.019 min
Response: 349613
Conc: 140.29 ng/ml



#34 AR-1260-4


R.T.: 7.855 min
Delta R.T.: -0.015 min
Response: 315842
Conc: 136.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#34 AR-1260-4

R.T.: 6.922 min
Delta R.T.: -0.020 min
Response: 238711
Conc: 115.68 ng/ml




#35 AR-1260-5

R.T.: 8.164 min
Delta R.T.: -0.015 min
Response: 525109
Conc: 123.78 ng/ml



#35 AR-1260-5


R.T.: 7.164 min
Delta R.T.: -0.018 min
Response: 554594
Conc: 110.43 ng/ml



#36 AR-1262-1


R.T.: 7.630 min
Delta R.T.: -0.015 min
Response: 258944
Conc: 80.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#36 AR-1262-1

R.T.: 6.663 min
Delta R.T.: -0.019 min
Response: 278074
Conc: 67.81 ng/ml




#37 AR-1262-2

R.T.: 8.164 min
Delta R.T.: -0.015 min
Response: 525109
Conc: 97.81 ng/ml



#37 AR-1262-2


R.T.: 7.164 min
Delta R.T.: -0.017 min
Response: 554594
Conc: 88.69 ng/ml



#38 AR-1262-3


R.T.: 8.474 min
Delta R.T.: 0.002 min
Response: 388628
Conc: 107.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#38 AR-1262-3

R.T.: 7.508 min
Delta R.T.: 0.046 min
Response: 387051
Conc: 157.15 ng/ml




#39 AR-1262-4

R.T.: 8.523 min
Delta R.T.: -0.034 min
Response: 208133
Conc: 76.66 ng/ml



#39 AR-1262-4


R.T.: 7.508 min
Delta R.T.: -0.018 min
Response: 387051
Conc: 94.52 ng/ml



#40 AR-1262-5


R.T.: 9.162 min
Delta R.T.: -0.019 min
Response: 119747
Conc: 66.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T. : 8.018 min
Response: 0
Conc: N.D.




#41 AR-1268-1

R.T.: 8.474 min
Delta R.T.: 0.003 min
Response: 388628
Conc: 64.35 ng/ml



#41 AR-1268-1


R.T.: 7.508 min
Delta R.T.: 0.047 min
Response: 387051
Conc: 55.86 ng/ml



#42 AR-1268-2


R.T.: 8.523 min
Delta R.T.: -0.038 min
Response: 208133
Conc: 38.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS




#42 AR-1268-2

R.T.: 7.508 min
Delta R.T.: -0.019 min
Response: 387051
Conc: 68.03 ng/ml




#44 AR-1268-4

R.T.: 9.162 min
Delta R.T.: -0.018 min
Response: 119747
Conc: 61.03 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 8.017 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.544 min
Delta R.T.: -0.021 min
Response: 46819
Conc: 3.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
CULVER-CBTC-WC1-051019MS



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 8.298 min
Response: 0
Conc: N.D.