

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051623\
 Data File : P0095123.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2023 17:56
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 16 18:11:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.416	3.615	161.3E6	68356408	46.829	51.086
2) SA Decachlor...	10.264	8.630	115.7E6	51598794	47.683	45.842
Target Compounds						
3) L1 AR-1016-1	5.595	4.697	45798487	20216462	449.587	509.723
4) L1 AR-1016-2	5.617	4.714	70052010	31315047	462.982	529.720
5) L1 AR-1016-3	5.680	4.890	44734373	16088862	451.204	512.520
6) L1 AR-1016-4	5.779	4.933	35164959	14333666	472.448	495.694
7) L1 AR-1016-5	6.077	5.145	38684965	17929629	463.667	489.901
31) L7 AR-1260-1	7.217	6.180	69955737	32718083	471.815	464.744
32) L7 AR-1260-2	7.479	6.370	78665411	37582305	485.097	470.111
33) L7 AR-1260-3	7.842	6.523	60503175	34653369	462.798	460.799
34) L7 AR-1260-4	8.070	6.996	62665430	28559633	448.146	456.229
35) L7 AR-1260-5	8.398	7.241	113.6E6	59942940	464.154	475.592



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051623\
Data File : P0095123.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2023 17:56
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 16 18:11:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.416 min
Delta R.T.: 0.000 min
Response: 161306111
Conc: 46.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.615 min
Delta R.T.: -0.002 min
Response: 68356408
Conc: 51.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.264 min
Delta R.T.: 0.011 min
Response: 115718769
Conc: 47.68 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.630 min
Delta R.T.: 0.000 min
Response: 51598794
Conc: 45.84 ng/ml



#3 AR-1016-1


R.T.: 5.595 min
Delta R.T.: 0.001 min
Response: 45798487
Conc: 449.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.697 min
Delta R.T.: -0.003 min
Response: 20216462
Conc: 509.72 ng/ml




#4 AR-1016-2

R.T.: 5.617 min
Delta R.T.: 0.000 min
Response: 70052010
Conc: 462.98 ng/ml



#4 AR-1016-2


R.T.: 4.714 min
Delta R.T.: -0.003 min
Response: 31315047
Conc: 529.72 ng/ml



#5 AR-1016-3


R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 44734373
Conc: 451.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.890 min
Delta R.T.: -0.004 min
Response: 16088862
Conc: 512.52 ng/ml




#6 AR-1016-4

R.T.: 5.779 min
Delta R.T.: 0.000 min
Response: 35164959
Conc: 472.45 ng/ml



#6 AR-1016-4


R.T.: 4.933 min
Delta R.T.: -0.003 min
Response: 14333666
Conc: 495.69 ng/ml



#7 AR-1016-5


R.T.: 6.077 min
Delta R.T.: 0.001 min
Response: 38684965
Conc: 463.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.145 min
Delta R.T.: -0.004 min
Response: 17929629
Conc: 489.90 ng/ml




#31 AR-1260-1

R.T.: 7.217 min
Delta R.T.: 0.003 min
Response: 69955737
Conc: 471.82 ng/ml



#31 AR-1260-1


R.T.: 6.180 min
Delta R.T.: -0.003 min
Response: 32718083
Conc: 464.74 ng/ml



#32 AR-1260-2


R.T.: 7.479 min
Delta R.T.: 0.004 min
Response: 78665411
Conc: 485.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.370 min
Delta R.T.: -0.003 min
Response: 37582305
Conc: 470.11 ng/ml




#33 AR-1260-3

R.T.: 7.842 min
Delta R.T.: 0.003 min
Response: 60503175
Conc: 462.80 ng/ml



#33 AR-1260-3


R.T.: 6.523 min
Delta R.T.: -0.003 min
Response: 34653369
Conc: 460.80 ng/ml



#34 AR-1260-4


R.T.: 8.070 min
Delta R.T.: 0.003 min
Response: 62665430
Conc: 448.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.996 min
Delta R.T.: -0.003 min
Response: 28559633
Conc: 456.23 ng/ml



#35 AR-1260-5

R.T.: 8.398 min
Delta R.T.: 0.004 min
Response: 11355542
Conc: 464.15 ng/ml



#35 AR-1260-5

R.T.: 7.241 min
Delta R.T.: -0.002 min
Response: 59942940
Conc: 475.59 ng/ml