

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051623\
 Data File : P0095124.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2023 18:13
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 16 18:37:18 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.415	3.616	171.0E6	70973080	49.639	53.042
2)	SA Decachlor...	10.263	8.629	116.6E6	53272172	48.055	47.329
Target Compounds							
16)	L4 AR-1242-1	5.595	4.697	34294601	17441902	410.532	526.659 #
17)	L4 AR-1242-2	5.617	4.714	52454290	27310713	417.241	554.033 #
18)	L4 AR-1242-3	5.680	4.890	33276918	14282344	408.635	548.923 #
19)	L4 AR-1242-4	5.779	4.974	26475309	15546933	431.090	525.216
20)	L4 AR-1242-5	6.525	5.497	34662992	16739915	538.636	529.827



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051623\
Data File : P0095124.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2023 18:13
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 16 18:37:18 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.415 min
Delta R.T.: 0.000 min
Response: 170987184
Conc: 49.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.616 min
Delta R.T.: -0.002 min
Response: 70973080
Conc: 53.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.263 min
Delta R.T.: 0.011 min
Response: 116622182
Conc: 48.05 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.629 min
Delta R.T.: -0.002 min
Response: 53272172
Conc: 47.33 ng/ml



#16 AR-1242-1


R.T.: 5.595 min
Delta R.T.: 0.000 min
Response: 34294601
Conc: 410.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.697 min
Delta R.T.: -0.004 min
Response: 17441902
Conc: 526.66 ng/ml




#17 AR-1242-2

R.T.: 5.617 min
Delta R.T.: 0.000 min
Response: 52454290
Conc: 417.24 ng/ml



#17 AR-1242-2


R.T.: 4.714 min
Delta R.T.: -0.004 min
Response: 27310713
Conc: 554.03 ng/ml



#18 AR-1242-3


R.T.: 5.680 min
Delta R.T.: 0.000 min
Response: 33276918
Conc: 408.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.890 min
Delta R.T.: -0.005 min
Response: 14282344
Conc: 548.92 ng/ml




#19 AR-1242-4

R.T.: 5.779 min
Delta R.T.: 0.000 min
Response: 26475309
Conc: 431.09 ng/ml



#19 AR-1242-4


R.T.: 4.974 min
Delta R.T.: -0.004 min
Response: 15546933
Conc: 525.22 ng/ml



#20 AR-1242-5

R.T.: 6.525 min
Delta R.T.: 0.001 min
Response: 34662992
Conc: 538.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.497 min
Delta R.T.: -0.004 min
Response: 16739915
Conc: 529.83 ng/ml