

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051624\
 Data File : P0103838.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2024 14:48
 Operator : YP/AJ
 Sample : P2492-01 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC247601-1-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 16 15:58:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 16 04:14:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.449	3.583	10132960	4247840	1.317m	1.686 #
2) SA Decachlor...	10.210	8.540	8189650	1838824	1.728	1.414m

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051624\
Data File : PO103838.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2024 14:48
Operator : YP/AJ
Sample : P2492-01 10X
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BC247601-1-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 16 15:58:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050624.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 16 04:14:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: -0.002 min
Response: 10132960
Conc: 1.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC247601-1-1




#1 Tetrachloro-m-xylene

R.T.: 3.583 min
Delta R.T.: -0.008 min
Response: 4247840
Conc: 1.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: 0.008 min
Response: 8189650
Conc: 1.73 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.540 min
Delta R.T.: -0.015 min
Response: 1838824
Conc: 1.41 ng/ml m