

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051624\
 Data File : P0103840.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2024 16:56
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 17 01:38:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 16 04:14:30 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.449	3.583	420.4E6	141.0E6	54.633	55.975
2) SA Decachlor...	10.208	8.539	253.7E6	60563700	53.523	46.556
Target Compounds						
16) L4 AR-1242-1	5.618	4.659	102.3E6	32977970	499.710	498.846m
17) L4 AR-1242-2	5.641	4.678	146.7E6	51973345	503.505	522.124
18) L4 AR-1242-3	5.703	4.852	86172585	28278593	497.146	516.683
19) L4 AR-1242-4	5.801	4.935	71137030	25193851	495.244	507.176
20) L4 AR-1242-5	6.537	5.452	70644646	30582545	493.374	500.257



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051624\
Data File : PO103840.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2024 16:56
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 17 01:38:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050624.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 16 04:14:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: -0.002 min
Response: 420440577
Conc: 54.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.583 min
Delta R.T.: -0.009 min
Response: 141012246
Conc: 55.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.208 min
Delta R.T.: 0.006 min
Response: 253712240
Conc: 53.52 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.539 min
Delta R.T.: -0.017 min
Response: 60563700
Conc: 46.56 ng/ml



#16 AR-1242-1


R.T.: 5.618 min
Delta R.T.: 0.000 min
Response: 102253845
Conc: 499.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.659 min
Delta R.T.: -0.011 min
Response: 32977970
Conc: 498.85 ng/ml m




#17 AR-1242-2

R.T.: 5.641 min
Delta R.T.: 0.000 min
Response: 146672097
Conc: 503.51 ng/ml



#17 AR-1242-2


R.T.: 4.678 min
Delta R.T.: -0.011 min
Response: 51973345
Conc: 522.12 ng/ml



#18 AR-1242-3


R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 86172585
Conc: 497.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.852 min
Delta R.T.: -0.011 min
Response: 28278593
Conc: 516.68 ng/ml




#19 AR-1242-4

R.T.: 5.801 min
Delta R.T.: 0.000 min
Response: 71137030
Conc: 495.24 ng/ml



#19 AR-1242-4


R.T.: 4.935 min
Delta R.T.: -0.011 min
Response: 25193851
Conc: 507.18 ng/ml



#20 AR-1242-5

R.T.: 6.537 min
Delta R.T.: 0.002 min
Response: 70644646
Conc: 493.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.452 min
Delta R.T.: -0.011 min
Response: 30582545
Conc: 500.26 ng/ml