

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051721\
 Data File : P0077873.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 17 May 2021 13:15
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 17 16:07:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 14 05:52:29 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.351	3.630	1361627	602362	23.196	23.673
2)	SA Decachlor...	9.968	8.824	899326	550373	21.754	22.901
Target Compounds							
9)	L2 AR-1221-2	0.000	3.973	0	8984	N.D.	37.634 #
28)	L6 AR-1254-3	7.158	0.000	26688	0	8.096	N.D. #
32)	L7 AR-1260-2	0.000	6.623	0	4978	N.D.	2.870 #
33)	L7 AR-1260-3	0.000	6.777	0	10236	N.D.	6.251 #
35)	L7 AR-1260-5	0.000	7.496	0	3977	N.D.	1.399 #
37)	L8 AR-1262-2	0.000	7.496	0	3977	N.D.	1.194 #
38)	L8 AR-1262-3	0.000	7.797	0	3153	N.D.	2.285 #
39)	L8 AR-1262-4	8.809f	0.000	1942	0	1.238	N.D. #
41)	L9 AR-1268-1	0.000	7.797	0	3153	N.D.	0.868 #
42)	L9 AR-1268-2	8.809f	0.000	1942	0	0.352	N.D. #
45)	L9 AR-1268-5	0.000	8.607	0	126518	N.D.	14.734 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051721\
Data File : P0077873.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 17 May 2021 13:15
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 17 16:07:15 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051321.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 14 05:52:29 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.351 min
Delta R.T.: -0.005 min
Response: 1361627
Conc: 23.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.630 min
Delta R.T.: -0.005 min
Response: 602362
Conc: 23.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.968 min
Delta R.T.: -0.010 min
Response: 899326
Conc: 21.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.824 min
Delta R.T.: -0.008 min
Response: 550373
Conc: 22.90 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 4.701 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK




#9 AR-1221-2

R.T.: 3.973 min
Delta R.T.: -0.010 min
Response: 8984
Conc: 37.63 ng/ml




#28 AR-1254-3

R.T.: 7.158 min
Delta R.T.: 0.012 min
Response: 26688
Conc: 8.10 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T.: 6.294 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 7.579 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK




#32 AR-1260-2

R.T.: 6.623 min
Delta R.T.: -0.005 min
Response: 4978
Conc: 2.87 ng/ml




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T.: 7.939 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 6.777 min
Delta R.T.: -0.002 min
Response: 10236
Conc: 6.25 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 8.478 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK




#35 AR-1260-5

R.T.: 7.496 min
Delta R.T.: -0.009 min
Response: 3977
Conc: 1.40 ng/ml




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T.: 8.478 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 7.496 min
Delta R.T.: -0.006 min
Response: 3977
Conc: 1.19 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T.: 8.751 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK




#38 AR-1262-3

R.T.: 7.797 min
Delta R.T.: 0.012 min
Response: 3153
Conc: 2.28 ng/ml




#39 AR-1262-4

R.T.: 8.809 min
Delta R.T.: -0.019 min
Response: 1942
Conc: 1.24 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 7.847 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 8.750 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK




#41 AR-1268-1

R.T.: 7.797 min
Delta R.T.: 0.012 min
Response: 3153
Conc: 0.87 ng/ml




#42 AR-1268-2

R.T.: 8.809 min
Delta R.T.: -0.021 min
Response: 1942
Conc: 0.35 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T.: 7.849 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 9.711 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK



#45 AR-1268-5

R.T.: 8.607 min
Delta R.T.: -0.005 min
Response: 126518
Conc: 14.73 ng/ml