

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0051721\  
 Data File : P0077888.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 17 May 2021 17:51  
 Operator : DD\AJ  
 Sample : M2409-01  
 Misc :  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 BACKFILL-ASHCROFT-REAR-06

Manual Integrations  
 APPROVED

Ankita  
 5/18/2021 4:19:36 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: May 18 04:38:37 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0051321.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Fri May 14 05:52:29 2021  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1  | RT#2  | Resp#1  | Resp#2 | ng/ml  | ng/ml   |
|-----------------------------|-------|-------|---------|--------|--------|---------|
| -----                       |       |       |         |        |        |         |
| System Monitoring Compounds |       |       |         |        |        |         |
| 1) SA Tetrachlo...          | 4.349 | 3.629 | 1452867 | 621636 | 24.750 | 24.430m |
| 2) SA Decachlor...          | 9.965 | 8.823 | 1087170 | 609267 | 26.298 | 25.352  |

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0051721\  
Data File : P0077888.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 17 May 2021 17:51  
Operator : DD\AJ  
Sample : M2409-01  
Misc :  
ALS Vial : 26 Sample Multiplier: 1



Instrument :  
ECD\_O  
ClientSampleId :  
BACKFILL-ASHCROFT-REAR-06


Manual Integrations  
APPROVED

Ankita  
5/18/2021 4:19:36 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: May 18 04:38:37 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0051321.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Fri May 14 05:52:29 2021  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.349 min  
Delta R.T.: -0.007 min  
Response: 1452867  
Conc: 24.75 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
BACKFILL-ASHCROFT-REAR-06


Manual Integrations  
APPROVED

Ankita  
5/18/2021 4:19:36 PM




#1 Tetrachloro-m-xylene

R.T.: 3.629 min  
Delta R.T.: -0.005 min  
Response: 621636  
Conc: 24.43 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.965 min  
Delta R.T.: -0.013 min  
Response: 1087170  
Conc: 26.30 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.823 min  
Delta R.T.: -0.010 min  
Response: 609267  
Conc: 25.35 ng/ml