

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051724\
Data File : P0103926.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 May 2024 22:08
Operator : YP/AJ
Sample : P2513-07
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-15

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 18 00:13:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050624.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 16 04:14:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.449	3.582	179.4E6	60409914	23.315	23.980
2) SA Decachlor...	10.204	8.535	114.2E6	26671960	24.095	20.503

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0051724\
Data File : P0103926.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 May 2024 22:08
Operator : YP/AJ
Sample : P2513-07
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-15

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 18 00:13:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0050624.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 16 04:14:30 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.449 min
Delta R.T.: -0.002 min
Response: 179422946
Conc: 23.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-15




#1 Tetrachloro-m-xylene

R.T.: 3.582 min
Delta R.T.: -0.009 min
Response: 60409914
Conc: 23.98 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.204 min
Delta R.T.: 0.002 min
Response: 114215055
Conc: 24.09 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.535 min
Delta R.T.: -0.021 min
Response: 26671960
Conc: 20.50 ng/ml