

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051725\
 Data File : P0111139.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2025 14:33
 Operator : YP/AJ
 Sample : Q2032-05MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-24MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/19/2025
 Supervised By :mohammad ahmed 05/20/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 16 15:10:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 15 07:11:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.683	3.679	118.2E6	101.6E6	20.094	18.640
2)	SA Decachlor...	8.725	8.675	92867826	34397357	17.988	18.592
Target Compounds							
3)	L1 AR-1016-1	4.772	4.758	117.1E6	98397119	520.442m	515.364m
4)	L1 AR-1016-2	4.791	4.776	163.2E6	135.9E6	511.723m	490.336m
5)	L1 AR-1016-3	4.848	4.952	113.9E6	76445117	524.725m	523.328
6)	L1 AR-1016-4	4.968	4.994	100.9E6	59698534	579.552	497.120
7)	L1 AR-1016-5	5.226	5.207	92248504	76385951	524.165	494.726
31)	L7 AR-1260-1	6.265	6.238	167.9E6	130.9E6	550.610	523.978
32)	L7 AR-1260-2	6.455	6.426	196.1E6	147.0E6	512.067	499.421
33)	L7 AR-1260-3	6.822	6.578	148.5E6	137.0E6	426.110	477.676
34)	L7 AR-1260-4	7.082	7.050	130.2E6	89503279	464.993	445.815
35)	L7 AR-1260-5	7.324	7.290	318.8E6	194.7E6	442.799	423.666

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051725\
Data File : PO111139.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2025 14:33
Operator : YP/AJ
Sample : Q2032-05MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-24MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/19/2025
Supervised By :mohammad ahmed 05/20/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 16 15:10:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051525.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 15 07:11:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

