

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051822\
 Data File : P0086125.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 May 2022 10:11
 Operator : YP\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 18 14:36:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 12 05:35:26 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.462	0.000	1705	0	0.017	N.D. #
Target Compounds							
3)	L1 AR-1016-1	5.593f	0.000	17441	0	5.457	N.D. #
10)	L2 AR-1221-3	4.905f	0.000	3324	0	1.233	N.D. #
11)	L3 AR-1232-1	4.905f	0.000	3324	0	1.624	N.D. #
16)	L4 AR-1242-1	5.593f	0.000	17441	0	6.416	N.D. #
20)	L4 AR-1242-5	6.554	5.615	430	9732	0.235	8.558 #
21)	L5 AR-1248-1	5.593f	0.000	17441	0	8.635	N.D. #
24)	L5 AR-1248-4	6.554	0.000	430	0	0.134	N.D. #
25)	L5 AR-1248-5	6.554	5.615	430	9732	0.138	6.195 #
26)	L6 AR-1254-1	6.525	5.615	34029	9732	10.112	3.864 #
28)	L6 AR-1254-3	7.110	0.000	6113	0	1.210	N.D. #
29)	L6 AR-1254-4	7.365	0.000	2327	0	0.625	N.D. #
35)	L7 AR-1260-5	0.000	7.390f	0	884	N.D.	0.204 #
37)	L8 AR-1262-2	0.000	7.390f	0	884	N.D.	0.161 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051822\
Data File : P0086125.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 May 2022 10:11
Operator : YP\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 18 14:36:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 12 05:35:26 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.462 min
Delta R.T.: -0.016 min
Response: 1705
Conc: 0.02 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T. : 3.684 min
Response: 0
Conc: N.D.




#2 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 10.327 min
Response: 0
Conc: N.D.



#2 Decachlorobiphenyl


R.T.: 0.000 min
Exp R.T. : 8.759 min
Response: 0
Conc: N.D.



#3 AR-1016-1


R.T.: 5.593 min
Delta R.T.: -0.067 min
Response: 17441
Conc: 5.46 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.792 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 4.905 min
Delta R.T.: 0.044 min
Response: 3324
Conc: 1.23 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 4.070 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.905 min
Delta R.T.: 0.052 min
Response: 3324
Conc: 1.62 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 4.069 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.593 min
Delta R.T.: -0.068 min
Response: 17441
Conc: 6.42 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.792 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 6.554 min
Delta R.T.: -0.033 min
Response: 430
Conc: 0.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#20 AR-1242-5

R.T.: 5.615 min
Delta R.T.: 0.015 min
Response: 9732
Conc: 8.56 ng/ml




#21 AR-1248-1

R.T.: 5.593 min
Delta R.T.: -0.067 min
Response: 17441
Conc: 8.63 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T. : 4.791 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 6.554 min
Delta R.T.: 0.006 min
Response: 430
Conc: 0.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.246 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 6.554 min
Delta R.T.: -0.034 min
Response: 430
Conc: 0.14 ng/ml



#25 AR-1248-5


R.T.: 5.615 min
Delta R.T.: -0.026 min
Response: 9732
Conc: 6.19 ng/ml



#26 AR-1254-1


R.T.: 6.525 min
Delta R.T.: 0.002 min
Response: 34029
Conc: 10.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 5.615 min
Delta R.T.: 0.015 min
Response: 9732
Conc: 3.86 ng/ml




#28 AR-1254-3

R.T.: 7.110 min
Delta R.T.: -0.001 min
Response: 6113
Conc: 1.21 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.154 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.365 min
Delta R.T.: -0.033 min
Response: 2327
Conc: 0.62 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.383 min
Response: 0
Conc: N.D.




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 8.453 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 7.390 min
Delta R.T.: 0.043 min
Response: 884
Conc: 0.20 ng/ml



#37 AR-1262-2

R.T.: 0.000 min
Exp R.T.: 8.450 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#37 AR-1262-2

R.T.: 7.390 min
Delta R.T.: 0.047 min
Response: 884
Conc: 0.16 ng/ml