

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051823\
 Data File : P0095154.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 May 2023 23:21
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 04:06:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 19 04:06:09 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.493	3.650	76218484	31039234	25.541	25.615
2) SA Decachlor...	10.551	8.712	62095532	26109398	25.848	27.178
Target Compounds						
16) L4 AR-1242-1	5.696	4.743	21371991	8431019	267.613	274.528
17) L4 AR-1242-2	5.719	4.763	30016491	11512940	263.342	269.474
18) L4 AR-1242-3	5.783	4.940	19434017	6374841	266.945	269.794
19) L4 AR-1242-4	5.884	5.025	15137211	6731486	260.992	271.060
20) L4 AR-1242-5	6.645	5.551	15468786	8114260	262.066	270.957



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051823\
Data File : P0095154.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 May 2023 23:21
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 04:06:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051823.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 19 04:06:09 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.493 min
Delta R.T.: 0.000 min
Response: 76218484
Conc: 25.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.650 min
Delta R.T.: 0.000 min
Response: 31039234
Conc: 25.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.551 min
Delta R.T.: 0.000 min
Response: 62095532
Conc: 25.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.712 min
Delta R.T.: 0.000 min
Response: 26109398
Conc: 27.18 ng/ml



#16 AR-1242-1


R.T.: 5.696 min
Delta R.T.: 0.000 min
Response: 21371991
Conc: 267.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 8431019
Conc: 274.53 ng/ml




#17 AR-1242-2

R.T.: 5.719 min
Delta R.T.: 0.000 min
Response: 30016491
Conc: 263.34 ng/ml



#17 AR-1242-2


R.T.: 4.763 min
Delta R.T.: 0.000 min
Response: 11512940
Conc: 269.47 ng/ml



#18 AR-1242-3


R.T.: 5.783 min
Delta R.T.: 0.000 min
Response: 19434017
Conc: 266.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.940 min
Delta R.T.: 0.000 min
Response: 6374841
Conc: 269.79 ng/ml




#19 AR-1242-4

R.T.: 5.884 min
Delta R.T.: 0.000 min
Response: 15137211
Conc: 260.99 ng/ml



#19 AR-1242-4


R.T.: 5.025 min
Delta R.T.: 0.000 min
Response: 6731486
Conc: 271.06 ng/ml



#20 AR-1242-5

R.T.: 6.645 min
Delta R.T.: 0.000 min
Response: 15468786
Conc: 262.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.551 min
Delta R.T.: 0.000 min
Response: 8114260
Conc: 270.96 ng/ml