

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051921\
 Data File : P0077987.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 May 2021 13:13
 Operator : DD\AJ
 Sample : M2435-16
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 02:24:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 19 06:28:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.344	3.623	1873684	798936	25.352	26.441
2)	SA Decachlor...	9.960	8.815	1122353	652859	22.701	22.452
Target Compounds							
8)	L2 AR-1221-1	4.590	0.000	41634	0	50.584	N.D. #
9)	L2 AR-1221-2	4.688	0.000	25692	0	41.805	N.D. #
20)	L4 AR-1242-5	6.608	0.000	10325	0	7.716	N.D. #
24)	L5 AR-1248-4	6.563	0.000	100150	0	43.026	N.D. #
25)	L5 AR-1248-5	6.608	0.000	10325	0	4.563	N.D. #
26)	L6 AR-1254-1	6.563f	0.000	100150	0	40.351	N.D. #
28)	L6 AR-1254-3	7.152f	0.000	33104	0	8.438	N.D. #
34)	L7 AR-1260-4	8.147	0.000	20747	0	7.745	N.D. #
35)	L7 AR-1260-5	8.464	0.000	17799	0	3.466	N.D. #
37)	L8 AR-1262-2	8.464	0.000	17799	0	2.957	N.D. #
39)	L8 AR-1262-4	8.820	0.000	21861	0	7.010	N.D. #
42)	L9 AR-1268-2	8.820	0.000	21861	0	3.378	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051921\
Data File : P0077987.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 May 2021 13:13
Operator : DD\AJ
Sample : M2435-16
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 02:24:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051821.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 19 06:28:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.344 min
Delta R.T.: 0.000 min
Response: 1873684
Conc: 25.35 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.623 min
Delta R.T.: 0.000 min
Response: 798936
Conc: 26.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.960 min
Delta R.T.: 0.005 min
Response: 1122353
Conc: 22.70 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.815 min
Delta R.T.: -0.001 min
Response: 652859
Conc: 22.45 ng/ml



#8 AR-1221-1


R.T.: 4.590 min
Delta R.T.: -0.007 min
Response: 41634
Conc: 50.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 3.878 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 4.688 min
Delta R.T.: -0.004 min
Response: 25692
Conc: 41.80 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 3.975 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 6.608 min
Delta R.T.: 0.005 min
Response: 10325
Conc: 7.72 ng/ml

Instrument :
ECD_O
ClientSampleId :




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 5.713 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.563 min
Delta R.T.: -0.002 min
Response: 100150
Conc: 43.03 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.338 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.608 min
Delta R.T.: 0.006 min
Response: 10325
Conc: 4.56 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 5.754 min
Response: 0
Conc: N.D.




#26 AR-1254-1

R.T.: 6.563 min
Delta R.T.: 0.029 min
Response: 100150
Conc: 40.35 ng/ml



#26 AR-1254-1


R.T.: 0.000 min
Exp R.T. : 5.712 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.152 min
Delta R.T.: 0.015 min
Response: 33104
Conc: 8.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.283 min
Response: 0
Conc: N.D.




#34 AR-1260-4

R.T.: 8.147 min
Delta R.T.: -0.004 min
Response: 20747
Conc: 7.75 ng/ml



#34 AR-1260-4


R.T.: 0.000 min
Exp R.T. : 7.242 min
Response: 0
Conc: N.D.



#35 AR-1260-5


R.T.: 8.464 min
Delta R.T.: 0.000 min
Response: 17799
Conc: 3.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.489 min
Response: 0
Conc: N.D.




#37 AR-1262-2

R.T.: 8.464 min
Delta R.T.: -0.002 min
Response: 17799
Conc: 2.96 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 7.491 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 8.820 min
Delta R.T.: 0.004 min
Response: 21861
Conc: 7.01 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 0.000 min
Exp R.T. : 7.837 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 8.820 min
Delta R.T.: 0.003 min
Response: 21861
Conc: 3.38 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 7.837 min
Response: 0
Conc: N.D.