

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051921\
 Data File : P0077994.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 May 2021 15:13
 Operator : DD\AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 02:27:05 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 19 06:28:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.344	3.624	3629173	1553219	49.104	51.404
2)	SA Decachlor...	9.959	8.816	2357780	1352516	47.689	46.514
Target Compounds							
21)	L5 AR-1248-1	5.640	4.854	697110	321059	489.233	498.196
22)	L5 AR-1248-2	5.931	5.113	935484	451886	492.749	472.712
23)	L5 AR-1248-3	6.142	5.156	1108113	460918	499.589	479.505
24)	L5 AR-1248-4	6.565	5.336	1200807	539167	515.890	470.781
25)	L5 AR-1248-5	6.602	5.753	1136765	519786	502.421	496.051



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051921\
Data File : P0077994.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 May 2021 15:13
Operator : DD\AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleID :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 02:27:05 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051821.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 19 06:28:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.344 min
Delta R.T.: 0.000 min
Response: 3629173
Conc: 49.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 1553219
Conc: 51.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.959 min
Delta R.T.: 0.004 min
Response: 2357780
Conc: 47.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.816 min
Delta R.T.: 0.000 min
Response: 1352516
Conc: 46.51 ng/ml



#21 AR-1248-1


R.T.: 5.640 min
Delta R.T.: 0.000 min
Response: 697110
Conc: 489.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.854 min
Delta R.T.: -0.001 min
Response: 321059
Conc: 498.20 ng/ml




#22 AR-1248-2

R.T.: 5.931 min
Delta R.T.: 0.000 min
Response: 935484
Conc: 492.75 ng/ml



#22 AR-1248-2


R.T.: 5.113 min
Delta R.T.: -0.001 min
Response: 451886
Conc: 472.71 ng/ml



#23 AR-1248-3


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 1108113
Conc: 499.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.156 min
Delta R.T.: -0.001 min
Response: 460918
Conc: 479.50 ng/ml




#24 AR-1248-4

R.T.: 6.565 min
Delta R.T.: 0.000 min
Response: 1200807
Conc: 515.89 ng/ml



#24 AR-1248-4


R.T.: 5.336 min
Delta R.T.: -0.002 min
Response: 539167
Conc: 470.78 ng/ml



#25 AR-1248-5

R.T.: 6.602 min
Delta R.T.: 0.000 min
Response: 1136765
Conc: 502.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.753 min
Delta R.T.: -0.001 min
Response: 519786
Conc: 496.05 ng/ml