

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051921\
 Data File : P0078007.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 19 May 2021 18:57
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Ankita
 5/20/2021 2:06:28 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 02:32:17 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 19 06:28:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.344	3.624	3643181	1586344	49.294	52.500
2)	SA Decachlor...	9.958	8.816	2096854	1196341	42.411	41.143
Target Compounds							
3)	L1 AR-1016-1	5.639	4.854	1103992	524504	453.205	480.388
4)	L1 AR-1016-2	5.661	4.873	1652714	753366	463.555	484.601
5)	L1 AR-1016-3	5.726	5.060	1051306	387567	474.432	473.498
6)	L1 AR-1016-4	5.830	5.113	779486	329752	446.565	463.576
7)	L1 AR-1016-5	6.141	5.337	726386	397166	419.887	450.517
31)	L7 AR-1260-1	7.301	6.417	1190548	689380	424.247m	430.254
32)	L7 AR-1260-2	7.566	6.614	1386074	841837	412.021	436.055
33)	L7 AR-1260-3	7.925	6.765	1049164	776070	429.988	423.011
34)	L7 AR-1260-4	8.151	7.242	1201124	638499	448.392	423.563
35)	L7 AR-1260-5	8.465	7.490	2277034	1475577	443.358	423.136

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051921\
Data File : P0078007.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 19 May 2021 18:57
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/20/2021 2:06:28 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 02:32:17 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051821.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 19 06:28:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.344 min
Delta R.T.: 0.000 min
Response: 3643181
Conc: 49.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/20/2021 2:06:28 PM




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 1586344
Conc: 52.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.958 min
Delta R.T.: 0.002 min
Response: 2096854
Conc: 42.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.816 min
Delta R.T.: 0.000 min
Response: 1196341
Conc: 41.14 ng/ml




#3 AR-1016-1

R.T.: 5.639 min
Delta R.T.: 0.000 min
Response: 1103992
Conc: 453.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/20/2021 2:06:28 PM




#3 AR-1016-1

R.T.: 4.854 min
Delta R.T.: 0.000 min
Response: 524504
Conc: 480.39 ng/ml




#4 AR-1016-2

R.T.: 5.661 min
Delta R.T.: 0.000 min
Response: 1652714
Conc: 463.55 ng/ml



#4 AR-1016-2

R.T.: 4.873 min
Delta R.T.: 0.000 min
Response: 753366
Conc: 484.60 ng/ml




#5 AR-1016-3

R.T.: 5.726 min
Delta R.T.: 0.000 min
Response: 1051306
Conc: 474.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/20/2021 2:06:28 PM




#5 AR-1016-3

R.T.: 5.060 min
Delta R.T.: 0.000 min
Response: 387567
Conc: 473.50 ng/ml




#6 AR-1016-4

R.T.: 5.830 min
Delta R.T.: 0.000 min
Response: 779486
Conc: 446.57 ng/ml



#6 AR-1016-4

R.T.: 5.113 min
Delta R.T.: 0.000 min
Response: 329752
Conc: 463.58 ng/ml




#7 AR-1016-5

R.T.: 6.141 min
Delta R.T.: 0.000 min
Response: 726386
Conc: 419.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/20/2021 2:06:28 PM




#7 AR-1016-5

R.T.: 5.337 min
Delta R.T.: 0.000 min
Response: 397166
Conc: 450.52 ng/ml




#31 AR-1260-1

R.T.: 7.301 min
Delta R.T.: 0.000 min
Response: 1190548
Conc: 424.25 ng/ml m



#31 AR-1260-1

R.T.: 6.417 min
Delta R.T.: 0.000 min
Response: 689380
Conc: 430.25 ng/ml




#32 AR-1260-2

R.T.: 7.566 min
Delta R.T.: 0.000 min
Response: 1386074
Conc: 412.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/20/2021 2:06:28 PM




#32 AR-1260-2

R.T.: 6.614 min
Delta R.T.: 0.000 min
Response: 841837
Conc: 436.05 ng/ml




#33 AR-1260-3

R.T.: 7.925 min
Delta R.T.: 0.001 min
Response: 1049164
Conc: 429.99 ng/ml



#33 AR-1260-3

R.T.: 6.765 min
Delta R.T.: 0.000 min
Response: 776070
Conc: 423.01 ng/ml




#34 AR-1260-4

R.T.: 8.151 min
Delta R.T.: 0.000 min
Response: 1201124
Conc: 448.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
5/20/2021 2:06:28 PM




#34 AR-1260-4

R.T.: 7.242 min
Delta R.T.: 0.000 min
Response: 638499
Conc: 423.56 ng/ml



#35 AR-1260-5

R.T.: 8.465 min
Delta R.T.: 0.001 min
Response: 2277034
Conc: 443.36 ng/ml



#35 AR-1260-5

R.T.: 7.490 min
Delta R.T.: 0.000 min
Response: 1475577
Conc: 423.14 ng/ml