

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051923\
 Data File : P0095180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2023 10:39
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 19 21:44:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 19 06:20:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.492	3.651	162.7E6	64320204	55.835	56.645
2) SA Decachlor...	10.542	8.710	134.6E6	52511275	58.018	54.145
Target Compounds						
21) L5 AR-1248-1	5.694	4.743	33808621	11979552	472.659	436.622
22) L5 AR-1248-2	5.977	4.982	51568453	19011664	486.412	472.447
23) L5 AR-1248-3	6.187	5.024	55297287	19957306	481.970	474.549
24) L5 AR-1248-4	6.603	5.197	60917605	22673830	483.204	470.630
25) L5 AR-1248-5	6.643	5.592	58976773	21552540	483.640	481.333



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051923\
Data File : PO095180.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2023 10:39
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 19 21:44:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051823.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 19 06:20:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.492 min
Delta R.T.: -0.002 min
Response: 162693693
Conc: 55.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.651 min
Delta R.T.: -0.001 min
Response: 64320204
Conc: 56.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.542 min
Delta R.T.: -0.006 min
Response: 134630895
Conc: 58.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.710 min
Delta R.T.: -0.003 min
Response: 52511275
Conc: 54.15 ng/ml



#21 AR-1248-1


R.T.: 5.694 min
Delta R.T.: -0.001 min
Response: 33808621
Conc: 472.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 11979552
Conc: 436.62 ng/ml




#22 AR-1248-2

R.T.: 5.977 min
Delta R.T.: -0.002 min
Response: 51568453
Conc: 486.41 ng/ml



#22 AR-1248-2


R.T.: 4.982 min
Delta R.T.: 0.000 min
Response: 19011664
Conc: 472.45 ng/ml



#23 AR-1248-3


R.T.: 6.187 min
Delta R.T.: -0.002 min
Response: 55297287
Conc: 481.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.024 min
Delta R.T.: 0.000 min
Response: 19957306
Conc: 474.55 ng/ml




#24 AR-1248-4

R.T.: 6.603 min
Delta R.T.: -0.003 min
Response: 60917605
Conc: 483.20 ng/ml



#24 AR-1248-4


R.T.: 5.197 min
Delta R.T.: 0.000 min
Response: 22673830
Conc: 470.63 ng/ml



#25 AR-1248-5

R.T.: 6.643 min
Delta R.T.: -0.003 min
Response: 58976773
Conc: 483.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.592 min
Delta R.T.: 0.000 min
Response: 21552540
Conc: 481.33 ng/ml