

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0052022\
 Data File : P0086288.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 May 2022 2:56
 Operator : YP\AJ
 Sample : N2872-15
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 P022-SS002-0006-01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/21/2022
 Supervised By :Ankita Jodhani 05/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 21 04:52:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 09:06:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.479	3.676	2251214	1023390	22.156	23.495
2) SA Decachlor...	10.334	8.732	1180568	637710	21.742	18.325
Target Compounds						
31) L7 AR-1260-1	7.280	6.272	137625	82829	45.475	42.557
32) L7 AR-1260-2	7.539	6.460	361141	169105	99.832	71.950 #
33) L7 AR-1260-3	7.901	6.614	345477	100171	125.173	46.585 #
34) L7 AR-1260-4	8.129	7.087	396034	154433	117.379	85.766m#
35) L7 AR-1260-5	8.457	7.329	752375	427209	121.927	98.621

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052022\
Data File : P0086288.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 May 2022 2:56
Operator : YP\AJ
Sample : N2872-15
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
P022-SS002-0006-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/21/2022
Supervised By :Ankita Jodhani 05/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 21 04:52:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.479 min
Delta R.T.: 0.001 min
Response: 2251214
Conc: 22.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
P022-SS002-0006-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/21/2022
Supervised By :Ankita Jodhani 05/23/2022




#1 Tetrachloro-m-xylene

R.T.: 3.676 min
Delta R.T.: -0.007 min
Response: 1023390
Conc: 23.49 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.334 min
Delta R.T.: 0.009 min
Response: 1180568
Conc: 21.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: -0.022 min
Response: 637710
Conc: 18.32 ng/ml




#31 AR-1260-1

R.T.: 7.280 min
Delta R.T.: 0.003 min
Response: 137625
Conc: 45.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
P022-SS002-0006-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/21/2022
Supervised By :Ankita Jodhani 05/23/2022




#31 AR-1260-1

R.T.: 6.272 min
Delta R.T.: -0.017 min
Response: 82829
Conc: 42.56 ng/ml




#32 AR-1260-2

R.T.: 7.539 min
Delta R.T.: 0.004 min
Response: 361141
Conc: 99.83 ng/ml



#32 AR-1260-2

R.T.: 6.460 min
Delta R.T.: -0.016 min
Response: 169105
Conc: 71.95 ng/ml




#33 AR-1260-3

R.T.: 7.901 min
Delta R.T.: 0.004 min
Response: 345477
Conc: 125.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
P022-SS002-0006-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/21/2022
Supervised By :Ankita Jodhani 05/23/2022




#33 AR-1260-3

R.T.: 6.614 min
Delta R.T.: -0.017 min
Response: 100171
Conc: 46.59 ng/ml




#34 AR-1260-4

R.T.: 8.129 min
Delta R.T.: 0.004 min
Response: 396034
Conc: 117.38 ng/ml



#34 AR-1260-4

R.T.: 7.087 min
Delta R.T.: -0.019 min
Response: 154433
Conc: 85.77 ng/ml m




#35 AR-1260-5

R.T.: 8.457 min
Delta R.T.: 0.004 min
Response: 752375
Conc: 121.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
P022-SS002-0006-01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/21/2022
Supervised By :Ankita Jodhani 05/23/2022



#35 AR-1260-5

R.T.: 7.329 min
Delta R.T.: -0.018 min
Response: 427209
Conc: 98.62 ng/ml