

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052122\
 Data File : P0086389.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 May 2022 14:48
 Operator : YP\AJ
 Sample : N2873-05MSD
 Misc :
 ALS Vial : 74 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 P031-SS001-0006-01MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/23/2022
 Supervised By :Ankita Jodhani 05/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 23 03:40:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 09:06:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.478	3.675	2672780	1012591	26.305	23.247
2) SA Decachlor...	10.334	8.729	1204355	752070	22.180	21.611
Target Compounds						
3) L1 AR-1016-1	5.661	4.777	2041986	905490	638.943	579.788m
4) L1 AR-1016-2	5.684	4.797	2885726	1244524	649.143	594.337
5) L1 AR-1016-3	5.747	4.974	1798450	661637	655.906	598.568
6) L1 AR-1016-4	5.847	5.017	1471988	524617	664.089	569.076
7) L1 AR-1016-5	6.144	5.232	1429602	675951	689.936	598.636
31) L7 AR-1260-1	7.280	6.270	2122672	1236169	701.382	635.132
32) L7 AR-1260-2	7.538	6.458	2575206	1509893	711.876	642.420
33) L7 AR-1260-3	7.900	6.612	1742810	1341736	631.453	623.985
34) L7 AR-1260-4	8.130	7.086	2173919	953489	644.318	529.530m
35) L7 AR-1260-5	8.457	7.326	3664058	2335045	593.781	539.045

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052122\
Data File : P0086389.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 May 2022 14:48
Operator : YP\AJ
Sample : N2873-05MSD
Misc :
ALS Vial : 74 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
P031-SS001-0006-01MSD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 03:40:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

