Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO052122\
Data File : PO086402.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 May 2022 19:55
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 03:44:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	4.482	3.676	5002715	2094035	49.236	48.075
2)	SA Decachlor...	10.335	8.731	2724318	1539606	50.172	44.241
	Target Compounds						
3)	L1 AR-1016-1	5.663	4.778	1718791	768592	537.814	492.131m
4)	L1 AR-1016-2	5.686	4.798	2401387	1026635	540.191	490.281
5)	L1 AR-1016-3	5.749	4.975	1519995	554240	554.352	501.408
6)	L1 AR-1016-4	5.848	5.017	1239807	456130	559.340	494.786
7)	L1 AR-1016-5	6.146	5.232	1184415	633436	571.607	560.983
31)	L7 AR-1260-1	7.282	6.270	1824955	994429	603.009	510.928
32)	L7 AR-1260-2	7.540	6.458	2001118	1211778	553.178	515.580
33)	L7 AR-1260-3	7.903	6.613	1580391	1092445	572.606	508.050
34)	L7 AR-1260-4	8.131	7.085	1854671	903304	549.697	501.659m
35)	L7 AR-1260-5	8.458	7.327	3193881	2131481	517.586	492.052

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO052122\
Data File : PO086402.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 May 2022 19:55
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 05/23/2022
Supervised By :Ankita Jodhani 05/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 03:44:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm