

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0051122\
 Data File : P0086337.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 21 May 2022 20:26
 Operator : YP\AJ
 Sample : N2872-12
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 P022-SS001-1218-01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/24/2022
 Supervised By :Ankita Jodhani 05/24/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 01:54:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 09:06:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.479	3.676	2294675	921014	22.584	21.145
2) SA Decachlor...	10.333	8.732	1003882	534020	18.488	15.345
Target Compounds						
31) L7 AR-1260-1	7.280	6.271	123209	64523	40.711	33.151
32) L7 AR-1260-2	7.539	6.460	818371	404660	226.226	172.172
33) L7 AR-1260-3	7.901	6.615	837875	155758	303.578	72.437 #
34) L7 AR-1260-4	8.128	7.086	756850	291935	224.319m	162.129m#
35) L7 AR-1260-5	8.457	7.328	2219793	1317818	359.730	304.218

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO052122\
Data File : PO086337.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 21 May 2022 20:26
Operator : YP\AJ
Sample : N2872-12
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
P022-SS001-1218-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 01:54:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 2294675
Conc: 22.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
P022-SS001-1218-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022




#1 Tetrachloro-m-xylene

R.T.: 3.676 min
Delta R.T.: -0.007 min
Response: 921014
Conc: 21.14 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.333 min
Delta R.T.: 0.008 min
Response: 1003882
Conc: 18.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: -0.022 min
Response: 534020
Conc: 15.35 ng/ml




#31 AR-1260-1

R.T.: 7.280 min
Delta R.T.: 0.003 min
Response: 123209
Conc: 40.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
P022-SS001-1218-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022




#31 AR-1260-1

R.T.: 6.271 min
Delta R.T.: -0.017 min
Response: 64523
Conc: 33.15 ng/ml




#32 AR-1260-2

R.T.: 7.539 min
Delta R.T.: 0.004 min
Response: 818371
Conc: 226.23 ng/ml



#32 AR-1260-2

R.T.: 6.460 min
Delta R.T.: -0.016 min
Response: 404660
Conc: 172.17 ng/ml




#33 AR-1260-3

R.T.: 7.901 min
Delta R.T.: 0.004 min
Response: 837875
Conc: 303.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
P022-SS001-1218-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022




#33 AR-1260-3

R.T.: 6.615 min
Delta R.T.: -0.017 min
Response: 155758
Conc: 72.44 ng/ml




#34 AR-1260-4

R.T.: 8.128 min
Delta R.T.: 0.004 min
Response: 756850
Conc: 224.32 ng/ml m



#34 AR-1260-4

R.T.: 7.086 min
Delta R.T.: -0.020 min
Response: 291935
Conc: 162.13 ng/ml m




#35 AR-1260-5

R.T.: 8.457 min
Delta R.T.: 0.004 min
Response: 2219793
Conc: 359.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
P022-SS001-1218-01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022



#35 AR-1260-5

R.T.: 7.328 min
Delta R.T.: -0.019 min
Response: 1317818
Conc: 304.22 ng/ml