

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052223\
 Data File : P0095217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 May 2023 10:03
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 22 21:34:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 19 06:20:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.494	3.650	155.9E6	59416807	53.514	52.327
2) SA Decachlor...	10.551	8.710	129.5E6	47689606	55.806	49.174
Target Compounds						
26) L6 AR-1254-1	6.580	5.551	60169256	29915391	532.022	482.505
27) L6 AR-1254-2	6.806	5.700	89592592	27106071	515.733	486.190
28) L6 AR-1254-3	7.184	6.107	91524603	41749117	521.724	486.705
29) L6 AR-1254-4	7.480	6.336	66944482	25137858	520.193	483.972
30) L6 AR-1254-5	7.916	6.758	87288430	38135776	526.696	493.132



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO052223\
Data File : PO095217.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 May 2023 10:03
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 22 21:34:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051823.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 19 06:20:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.494 min
Delta R.T.: 0.000 min
Response: 155929486
Conc: 53.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.650 min
Delta R.T.: -0.001 min
Response: 59416807
Conc: 52.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.551 min
Delta R.T.: 0.002 min
Response: 129496803
Conc: 55.81 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.710 min
Delta R.T.: -0.003 min
Response: 47689606
Conc: 49.17 ng/ml



#26 AR-1254-1


R.T.: 6.580 min
Delta R.T.: 0.000 min
Response: 60169256
Conc: 532.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.551 min
Delta R.T.: 0.000 min
Response: 29915391
Conc: 482.50 ng/ml




#27 AR-1254-2

R.T.: 6.806 min
Delta R.T.: 0.000 min
Response: 89592592
Conc: 515.73 ng/ml



#27 AR-1254-2


R.T.: 5.700 min
Delta R.T.: 0.000 min
Response: 27106071
Conc: 486.19 ng/ml



#28 AR-1254-3


R.T.: 7.184 min
Delta R.T.: 0.000 min
Response: 91524603
Conc: 521.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.107 min
Delta R.T.: 0.000 min
Response: 41749117
Conc: 486.71 ng/ml




#29 AR-1254-4

R.T.: 7.480 min
Delta R.T.: 0.000 min
Response: 66944482
Conc: 520.19 ng/ml



#29 AR-1254-4


R.T.: 6.336 min
Delta R.T.: 0.000 min
Response: 25137858
Conc: 483.97 ng/ml



#30 AR-1254-5

R.T.: 7.916 min
Delta R.T.: 0.000 min
Response: 87288430
Conc: 526.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.758 min
Delta R.T.: 0.000 min
Response: 38135776
Conc: 493.13 ng/ml