

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052322\
 Data File : P0086440.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 May 2022 22:34
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/24/2022
 Supervised By :Ankita Jodhani 05/24/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 01:45:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 09:06:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.480	3.676	4812225	2031564	47.361	46.641
2)	SA Decachlor...	10.331	8.727	2608737	1452550	48.044	41.739
Target Compounds							
3)	L1 AR-1016-1	5.662	4.777	1619871	726231	506.862	465.008m
4)	L1 AR-1016-2	5.684	4.797	2269577	977254	510.541	466.699
5)	L1 AR-1016-3	5.747	4.974	1426859	524349	520.384	474.366
6)	L1 AR-1016-4	5.847	5.016	1212039	430884	546.813	467.400
7)	L1 AR-1016-5	6.144	5.231	1082987	561902	522.657	497.631
31)	L7 AR-1260-1	7.280	6.268	1614989	918167	533.631	471.745
32)	L7 AR-1260-2	7.537	6.456	1889107	1105088	522.215	470.186
33)	L7 AR-1260-3	7.900	6.610	1376067	990983	498.575	460.865
34)	L7 AR-1260-4	8.128	7.084	1628282	818212	482.599	454.402m
35)	L7 AR-1260-5	8.455	7.325	2988918	1918854	484.371	442.967

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052322\
Data File : P0086440.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 May 2022 22:34
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 01:45:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

