

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052322\
 Data File : P0086443.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 May 2022 23:55
 Operator : YP\AJ
 Sample : N2876-05
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 P007-SS011-0612-01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/24/2022
 Supervised By :Ankita Jodhani 05/24/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 01:47:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 09:06:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.478	3.675	1725796	826452	16.985	18.974
2) SA Decachlor...	10.330	8.727	1197423	628145	22.052	18.050
Target Compounds						
31) L7 AR-1260-1	7.279	6.269	185.5E6	99265860	61302.967	51001.838
32) L7 AR-1260-2	7.538	6.457	237.8E6	123.8E6	65730.058	52689.432
33) L7 AR-1260-3	7.901	6.611	173.2E6	118.9E6	62740.096	55305.383
34) L7 AR-1260-4	8.128	7.083	230.1E6	106.3E6	68199.257	59051.709m
35) L7 AR-1260-5	8.456	7.325	491.5E6	231.7E6	79643.039	53492.132 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052322\
Data File : P0086443.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 May 2022 23:55
Operator : YP\AJ
Sample : N2876-05
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
P007-SS011-0612-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 01:47:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.478 min
Delta R.T.: 0.000 min
Response: 1725796
Conc: 16.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
P007-SS011-0612-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022




#1 Tetrachloro-m-xylene

R.T.: 3.675 min
Delta R.T.: -0.008 min
Response: 826452
Conc: 18.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.330 min
Delta R.T.: 0.005 min
Response: 1197423
Conc: 22.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.727 min
Delta R.T.: -0.027 min
Response: 628145
Conc: 18.05 ng/ml




#31 AR-1260-1

R.T.: 7.279 min
Delta R.T.: 0.002 min
Response: 185528241
Conc: 61302.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
P007-SS011-0612-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022




#31 AR-1260-1

R.T.: 6.269 min
Delta R.T.: -0.020 min
Response: 99265860
Conc: 51001.84 ng/ml




#32 AR-1260-2

R.T.: 7.538 min
Delta R.T.: 0.003 min
Response: 237777976
Conc: 65730.06 ng/ml



#32 AR-1260-2

R.T.: 6.457 min
Delta R.T.: -0.019 min
Response: 123837122
Conc: 52689.43 ng/ml




#33 AR-1260-3

R.T.: 7.901 min
Delta R.T.: 0.004 min
Response: 173162626
Conc: 62740.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
P007-SS011-0612-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022




#33 AR-1260-3

R.T.: 6.611 min
Delta R.T.: -0.021 min
Response: 118921527
Conc: 55305.38 ng/ml




#34 AR-1260-4

R.T.: 8.128 min
Delta R.T.: 0.004 min
Response: 230103253
Conc: 68199.26 ng/ml



#34 AR-1260-4

R.T.: 7.083 min
Delta R.T.: -0.023 min
Response: 106330512
Conc: 59051.71 ng/ml m




#35 AR-1260-5

R.T.: 8.456 min
Delta R.T.: 0.003 min
Response: 491455061
Conc: 79643.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
P007-SS011-0612-01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/24/2022
Supervised By :Ankita Jodhani 05/24/2022



#35 AR-1260-5

R.T.: 7.325 min
Delta R.T.: -0.021 min
Response: 231718217
Conc: 53492.13 ng/ml