

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052323\
 Data File : P0095249.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 May 2023 10:54
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 23 21:07:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 19 06:20:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.495	3.650	151.3E6	58712914	51.928	51.707
2) SA Decachlor...	10.553	8.707	124.1E6	45298050	53.480	46.708
Target Compounds						
16) L4 AR-1242-1	5.698	4.742	41048131	15002653	514.610	490.560
17) L4 AR-1242-2	5.721	4.761	59300539	20798658	518.793	489.778
18) L4 AR-1242-3	5.785	4.938	40179188	11913590	548.946	511.356
19) L4 AR-1242-4	5.886	5.023	33621744	12295980	578.978	499.343
20) L4 AR-1242-5	6.648	5.549	30874647	14955276	528.661	503.880



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052323\
Data File : P0095249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2023 10:54
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 21:07:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051823.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 19 06:20:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.495 min
Delta R.T.: 0.001 min
Response: 151309220
Conc: 51.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.650 min
Delta R.T.: -0.002 min
Response: 58712914
Conc: 51.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.553 min
Delta R.T.: 0.004 min
Response: 124100399
Conc: 53.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.707 min
Delta R.T.: -0.006 min
Response: 45298050
Conc: 46.71 ng/ml



#16 AR-1242-1


R.T.: 5.698 min
Delta R.T.: 0.003 min
Response: 41048131
Conc: 514.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.742 min
Delta R.T.: -0.002 min
Response: 15002653
Conc: 490.56 ng/ml




#17 AR-1242-2

R.T.: 5.721 min
Delta R.T.: 0.003 min
Response: 59300539
Conc: 518.79 ng/ml



#17 AR-1242-2


R.T.: 4.761 min
Delta R.T.: -0.002 min
Response: 20798658
Conc: 489.78 ng/ml



#18 AR-1242-3


R.T.: 5.785 min
Delta R.T.: 0.003 min
Response: 40179188
Conc: 548.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.938 min
Delta R.T.: -0.002 min
Response: 11913590
Conc: 511.36 ng/ml




#19 AR-1242-4

R.T.: 5.886 min
Delta R.T.: 0.003 min
Response: 33621744
Conc: 578.98 ng/ml



#19 AR-1242-4


R.T.: 5.023 min
Delta R.T.: -0.002 min
Response: 12295980
Conc: 499.34 ng/ml



#20 AR-1242-5

R.T.: 6.648 min
Delta R.T.: 0.003 min
Response: 30874647
Conc: 528.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.549 min
Delta R.T.: -0.002 min
Response: 14955276
Conc: 503.88 ng/ml