

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052323\
 Data File : P0095279.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 May 2023 22:50
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 01:22:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051823.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 19 06:20:17 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.494	3.649	155.1E6	58959280	53.226	51.924
2) SA Decachlor...	10.560	8.706	122.5E6	45375790	52.801	46.788
Target Compounds						
16) L4 AR-1242-1	5.698	4.741	40058259	15205928	502.200	497.206
17) L4 AR-1242-2	5.720	4.760	56846626	21151970	497.325	498.098
18) L4 AR-1242-3	5.784	4.937	36848120	12032966	503.435	516.480
19) L4 AR-1242-4	5.885	5.022	30054715	12350063	517.552	501.540
20) L4 AR-1242-5	6.647	5.548	30488393	15114615	522.048	509.249



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052323\
Data File : P0095279.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 May 2023 22:50
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 01:22:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051823.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 19 06:20:17 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.494 min
Delta R.T.: 0.000 min
Response: 155090760
Conc: 53.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.649 min
Delta R.T.: -0.002 min
Response: 58959280
Conc: 51.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.560 min
Delta R.T.: 0.011 min
Response: 122524611
Conc: 52.80 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.706 min
Delta R.T.: -0.008 min
Response: 45375790
Conc: 46.79 ng/ml



#16 AR-1242-1


R.T.: 5.698 min
Delta R.T.: 0.002 min
Response: 40058259
Conc: 502.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.741 min
Delta R.T.: -0.003 min
Response: 15205928
Conc: 497.21 ng/ml




#17 AR-1242-2

R.T.: 5.720 min
Delta R.T.: 0.002 min
Response: 56846626
Conc: 497.32 ng/ml



#17 AR-1242-2


R.T.: 4.760 min
Delta R.T.: -0.003 min
Response: 21151970
Conc: 498.10 ng/ml



#18 AR-1242-3


R.T.: 5.784 min
Delta R.T.: 0.002 min
Response: 36848120
Conc: 503.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.937 min
Delta R.T.: -0.003 min
Response: 12032966
Conc: 516.48 ng/ml




#19 AR-1242-4

R.T.: 5.885 min
Delta R.T.: 0.002 min
Response: 30054715
Conc: 517.55 ng/ml



#19 AR-1242-4


R.T.: 5.022 min
Delta R.T.: -0.003 min
Response: 12350063
Conc: 501.54 ng/ml



#20 AR-1242-5

R.T.: 6.647 min
Delta R.T.: 0.003 min
Response: 30488393
Conc: 522.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.548 min
Delta R.T.: -0.003 min
Response: 15114615
Conc: 509.25 ng/ml