

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0052422\
 Data File : P0086516.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 May 2022 00:38
 Operator : YP\AJ
 Sample : N2896-16
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 P039-SS001-2430-01

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/25/2022
 Supervised By :Ankita Jodhani 05/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 25 02:44:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 09:06:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.476	3.672	1834424	804253	18.054	18.464
2) SA Decachlor...	10.327	8.722	950159	550971	17.499	15.832
Target Compounds						
31) L7 AR-1260-1	7.277	6.265	628036	374366	207.518	192.346
32) L7 AR-1260-2	7.535	6.452	1159386	643633	320.495	273.849m
33) L7 AR-1260-3	7.897	6.609	1595480	415264	578.073	193.122 #
34) L7 AR-1260-4	8.125	7.079	1844318	876176	546.629m	486.593m
35) L7 AR-1260-5	8.453	7.321	4334675	2766699	702.458	638.692

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO052422\
Data File : PO086516.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 May 2022 00:38
Operator : YP\AJ
Sample : N2896-16
Misc :
ALS Vial : 50 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
P039-SS001-2430-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2022
Supervised By :Ankita Jodhani 05/25/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 02:44:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.476 min
Delta R.T.: -0.002 min
Response: 1834424
Conc: 18.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
P039-SS001-2430-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2022
Supervised By :Ankita Jodhani 05/25/2022




#1 Tetrachloro-m-xylene

R.T.: 3.672 min
Delta R.T.: -0.012 min
Response: 804253
Conc: 18.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.327 min
Delta R.T.: 0.002 min
Response: 950159
Conc: 17.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.722 min
Delta R.T.: -0.032 min
Response: 550971
Conc: 15.83 ng/ml




#31 AR-1260-1

R.T.: 7.277 min
Delta R.T.: 0.000 min
Response: 628036
Conc: 207.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
P039-SS001-2430-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2022
Supervised By :Ankita Jodhani 05/25/2022




#31 AR-1260-1

R.T.: 6.265 min
Delta R.T.: -0.024 min
Response: 374366
Conc: 192.35 ng/ml




#32 AR-1260-2

R.T.: 7.535 min
Delta R.T.: 0.000 min
Response: 1159386
Conc: 320.49 ng/ml



#32 AR-1260-2

R.T.: 6.452 min
Delta R.T.: -0.024 min
Response: 643633
Conc: 273.85 ng/ml m




#33 AR-1260-3

R.T.: 7.897 min
Delta R.T.: 0.000 min
Response: 1595480
Conc: 578.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
P039-SS001-2430-01


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2022
Supervised By :Ankita Jodhani 05/25/2022




#33 AR-1260-3

R.T.: 6.609 min
Delta R.T.: -0.022 min
Response: 415264
Conc: 193.12 ng/ml




#34 AR-1260-4

R.T.: 8.125 min
Delta R.T.: 0.000 min
Response: 1844318
Conc: 546.63 ng/ml m



#34 AR-1260-4

R.T.: 7.079 min
Delta R.T.: -0.027 min
Response: 876176
Conc: 486.59 ng/ml m




#35 AR-1260-5

R.T.: 8.453 min
Delta R.T.: 0.000 min
Response: 4334675
Conc: 702.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
P039-SS001-2430-01

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2022
Supervised By :Ankita Jodhani 05/25/2022



#35 AR-1260-5

R.T.: 7.321 min
Delta R.T.: -0.026 min
Response: 2766699
Conc: 638.69 ng/ml