

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052422\
 Data File : P0086525.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 May 2022 3:35
 Operator : YP\AJ
 Sample : N2878-17
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 P007-SS022-0612-02

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/25/2022
 Supervised By :Ankita Jodhani 05/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 25 06:31:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 09:06:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.478	3.674	2674579	1103014	26.323	25.323
2) SA Decachlor...	10.328	8.721	1233015	716166	22.708	20.579
Target Compounds						
31) L7 AR-1260-1	7.277	6.265	144315	75574	47.685	38.829
32) L7 AR-1260-2	7.535	6.453	494991	280029	136.833	119.145
33) L7 AR-1260-3	7.898	6.609	927239	99096	335.956	46.085 #
34) L7 AR-1260-4	8.124	7.079	471498	186539	139.745m	103.596m#
35) L7 AR-1260-5	8.453	7.321	1508731	958240	244.498	221.210

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052422\
Data File : P0086525.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 May 2022 3:35
Operator : YP\AJ
Sample : N2878-17
Misc :
ALS Vial : 57 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
P007-SS022-0612-02


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2022
Supervised By :Ankita Jodhani 05/25/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 06:31:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.478 min
Delta R.T.: 0.000 min
Response: 2674579
Conc: 26.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
P007-SS022-0612-02


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2022
Supervised By :Ankita Jodhani 05/25/2022




#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: -0.009 min
Response: 1103014
Conc: 25.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.328 min
Delta R.T.: 0.003 min
Response: 1233015
Conc: 22.71 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.721 min
Delta R.T.: -0.032 min
Response: 716166
Conc: 20.58 ng/ml




#31 AR-1260-1

R.T.: 7.277 min
Delta R.T.: 0.000 min
Response: 144315
Conc: 47.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
P007-SS022-0612-02


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2022
Supervised By :Ankita Jodhani 05/25/2022




#31 AR-1260-1

R.T.: 6.265 min
Delta R.T.: -0.024 min
Response: 75574
Conc: 38.83 ng/ml




#32 AR-1260-2

R.T.: 7.535 min
Delta R.T.: 0.000 min
Response: 494991
Conc: 136.83 ng/ml



#32 AR-1260-2

R.T.: 6.453 min
Delta R.T.: -0.023 min
Response: 280029
Conc: 119.14 ng/ml




#33 AR-1260-3

R.T.: 7.898 min
Delta R.T.: 0.000 min
Response: 927239
Conc: 335.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
P007-SS022-0612-02


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2022
Supervised By :Ankita Jodhani 05/25/2022




#33 AR-1260-3

R.T.: 6.609 min
Delta R.T.: -0.022 min
Response: 99096
Conc: 46.09 ng/ml




#34 AR-1260-4

R.T.: 8.124 min
Delta R.T.: 0.000 min
Response: 471498
Conc: 139.75 ng/ml m



#34 AR-1260-4

R.T.: 7.079 min
Delta R.T.: -0.027 min
Response: 186539
Conc: 103.60 ng/ml m




#35 AR-1260-5

R.T.: 8.453 min
Delta R.T.: 0.000 min
Response: 1508731
Conc: 244.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
P007-SS022-0612-02

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/25/2022
Supervised By :Ankita Jodhani 05/25/2022



#35 AR-1260-5

R.T.: 7.321 min
Delta R.T.: -0.026 min
Response: 958240
Conc: 221.21 ng/ml