

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052424\
 Data File : P0104130.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 May 2024 13:00
 Operator : YP/AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 13:14:45 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 24 13:11:58 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.418	3.566	207.5E6	56005357	24.797	24.422
2) SA Decachlor...	10.130	8.502	118.7E6	51046200	25.582	25.463
Target Compounds						
21) L5 AR-1248-1	5.581	4.631	35286105	8113522	255.744	246.493
22) L5 AR-1248-2	5.854	4.865	55863782	13006629	257.847	258.418
23) L5 AR-1248-3	6.058	4.905	58035227	14295206	255.993	258.369
24) L5 AR-1248-4	6.461	5.075	54476838	17111418	251.883	263.855
25) L5 AR-1248-5	6.500	5.463	52855211	14180820	250.090	260.611



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO052424\
Data File : PO104130.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 May 2024 13:00
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 13:14:45 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO052424.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 24 13:11:58 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.418 min
Delta R.T.: 0.000 min
Response: 207492606
Conc: 24.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.566 min
Delta R.T.: 0.000 min
Response: 56005357
Conc: 24.42 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.130 min
Delta R.T.: -0.003 min
Response: 118656019
Conc: 25.58 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.502 min
Delta R.T.: 0.000 min
Response: 51046200
Conc: 25.46 ng/ml



#21 AR-1248-1


R.T.: 5.581 min
Delta R.T.: 0.000 min
Response: 35286105
Conc: 255.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 4.631 min
Delta R.T.: 0.000 min
Response: 8113522
Conc: 246.49 ng/ml




#22 AR-1248-2

R.T.: 5.854 min
Delta R.T.: 0.000 min
Response: 55863782
Conc: 257.85 ng/ml



#22 AR-1248-2


R.T.: 4.865 min
Delta R.T.: 0.000 min
Response: 13006629
Conc: 258.42 ng/ml



#23 AR-1248-3


R.T.: 6.058 min
Delta R.T.: 0.000 min
Response: 58035227
Conc: 255.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 4.905 min
Delta R.T.: 0.000 min
Response: 14295206
Conc: 258.37 ng/ml




#24 AR-1248-4

R.T.: 6.461 min
Delta R.T.: 0.000 min
Response: 54476838
Conc: 251.88 ng/ml



#24 AR-1248-4


R.T.: 5.075 min
Delta R.T.: 0.000 min
Response: 17111418
Conc: 263.85 ng/ml



#25 AR-1248-5

R.T.: 6.500 min
Delta R.T.: 0.000 min
Response: 52855211
Conc: 250.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 5.463 min
Delta R.T.: 0.000 min
Response: 14180820
Conc: 260.61 ng/ml