

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052722\
 Data File : P0086701.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 May 2022 13:30
 Operator : YP\AJ
 Sample : N3055-02 10X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 EDA7916-1-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 27 23:31:38 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 09:06:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.478	3.674	196166	85196	1.931	1.956
2) SA Decachlor...	10.330	8.724	121108	66575	2.230	1.913

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052722\
Data File : P0086701.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 May 2022 13:30
Operator : YP\AJ
Sample : N3055-02 10X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
EDA7916-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 27 23:31:38 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.478 min
Delta R.T.: 0.000 min
Response: 196166
Conc: 1.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
EDA7916-1-2




#1 Tetrachloro-m-xylene

R.T.: 3.674 min
Delta R.T.: -0.009 min
Response: 85196
Conc: 1.96 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.330 min
Delta R.T.: 0.005 min
Response: 121108
Conc: 2.23 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.724 min
Delta R.T.: -0.030 min
Response: 66575
Conc: 1.91 ng/ml