

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052722\
 Data File : P0086729.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 May 2022 22:31
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/28/2022
 Supervised By :Ankita Jodhani 05/28/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 00:01:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 19 09:06:34 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.478	3.672	5462679	2393114	53.763	54.941
2)	SA Decachlor...	10.330	8.723	3188354	1941792	58.718	55.798
Target Compounds							
26)	L6 AR-1254-1	6.524	5.581	1757167	1202563	522.132	477.454
27)	L6 AR-1254-2	6.744	5.729	2604339	1047060	529.225	478.146
28)	L6 AR-1254-3	7.113	6.134	2737798	1694503	541.880	479.610
29)	L6 AR-1254-4	7.401	6.362	2002151	1033021	537.531	466.963m
30)	L6 AR-1254-5	7.823	6.782	2055785	1453689	527.600	491.470

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO052722\
Data File : PO086729.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 May 2022 22:31
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/28/2022
Supervised By :Ankita Jodhani 05/28/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 00:01:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051122.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 19 09:06:34 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.478 min
Delta R.T.: 0.000 min
Response: 5462679
Conc: 53.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/28/2022
Supervised By :Ankita Jodhani 05/28/2022




#1 Tetrachloro-m-xylene

R.T.: 3.672 min
Delta R.T.: -0.011 min
Response: 2393114
Conc: 54.94 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.330 min
Delta R.T.: 0.005 min
Response: 3188354
Conc: 58.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.723 min
Delta R.T.: -0.031 min
Response: 1941792
Conc: 55.80 ng/ml




#26 AR-1254-1

R.T.: 6.524 min
Delta R.T.: 0.002 min
Response: 1757167
Conc: 522.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/28/2022
Supervised By :Ankita Jodhani 05/28/2022




#26 AR-1254-1

R.T.: 5.581 min
Delta R.T.: -0.019 min
Response: 1202563
Conc: 477.45 ng/ml




#27 AR-1254-2

R.T.: 6.744 min
Delta R.T.: 0.001 min
Response: 2604339
Conc: 529.23 ng/ml



#27 AR-1254-2

R.T.: 5.729 min
Delta R.T.: -0.019 min
Response: 1047060
Conc: 478.15 ng/ml




#28 AR-1254-3

R.T.: 7.113 min
Delta R.T.: 0.002 min
Response: 2737798
Conc: 541.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/28/2022
Supervised By :Ankita Jodhani 05/28/2022




#28 AR-1254-3

R.T.: 6.134 min
Delta R.T.: -0.020 min
Response: 1694503
Conc: 479.61 ng/ml




#29 AR-1254-4

R.T.: 7.401 min
Delta R.T.: 0.003 min
Response: 2002151
Conc: 537.53 ng/ml



#29 AR-1254-4

R.T.: 6.362 min
Delta R.T.: -0.021 min
Response: 1033021
Conc: 466.96 ng/ml m




#30 AR-1254-5

R.T.: 7.823 min
Delta R.T.: 0.003 min
Response: 2055785
Conc: 527.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/28/2022
Supervised By :Ankita Jodhani 05/28/2022



#30 AR-1254-5

R.T.: 6.782 min
Delta R.T.: -0.022 min
Response: 1453689
Conc: 491.47 ng/ml