

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052821\
 Data File : P0078252.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 May 2021 13:22
 Operator : DD\AJ
 Sample : M2551-05
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 JCSY-SP2

Manual Integrations
 APPROVED

Ankita
 6/4/2021 10:50:32 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 17:04:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 19 06:28:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.340	3.618	952788	490818	12.892m	16.244 #
2) SA Decachlor...	9.954	8.808	981838	638470	19.859	21.957
Target Compounds						
41) L9 AR-1268-1	8.733	7.765	1390189	662972	196.119	144.810 #
42) L9 AR-1268-2	8.814	7.829	1232988	1078622	190.535	260.694 #
43) L9 AR-1268-3	8.996	8.032	146610	130262	26.939	36.777 #
44) L9 AR-1268-4	9.359	8.323	801437	488019	355.865	322.561
45) L9 AR-1268-5	9.691	8.592	987885	538799	56.488	51.147

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO052821\
Data File : PO078252.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 28 May 2021 13:22
Operator : DD\AJ
Sample : M2551-05
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
JCSY-SP2


Manual Integrations
APPROVED

Ankita
6/4/2021 10:50:32 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 17:04:16 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051821.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 19 06:28:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.340 min
Delta R.T.: -0.004 min
Response: 952788
Conc: 12.89 ng/ml m

Instrument :
ECD_O
ClientSampleId :
JCSY-SP2


Manual Integrations
APPROVED

Ankita
6/4/2021 10:50:32 AM




#1 Tetrachloro-m-xylene

R.T.: 3.618 min
Delta R.T.: -0.005 min
Response: 490818
Conc: 16.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.954 min
Delta R.T.: -0.001 min
Response: 981838
Conc: 19.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.808 min
Delta R.T.: -0.008 min
Response: 638470
Conc: 21.96 ng/ml




#41 AR-1268-1

R.T.: 8.733 min
Delta R.T.: -0.002 min
Response: 1390189
Conc: 196.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
JCSY-SP2


Manual Integrations
APPROVED

Ankita
6/4/2021 10:50:32 AM




#41 AR-1268-1

R.T.: 7.765 min
Delta R.T.: -0.007 min
Response: 662972
Conc: 144.81 ng/ml




#42 AR-1268-2

R.T.: 8.814 min
Delta R.T.: -0.003 min
Response: 1232988
Conc: 190.54 ng/ml



#42 AR-1268-2

R.T.: 7.829 min
Delta R.T.: -0.008 min
Response: 1078622
Conc: 260.69 ng/ml




#43 AR-1268-3

R.T.: 8.996 min
Delta R.T.: -0.003 min
Response: 146610
Conc: 26.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
JCSY-SP2


Manual Integrations
APPROVED

Ankita
6/4/2021 10:50:32 AM




#43 AR-1268-3

R.T.: 8.032 min
Delta R.T.: -0.008 min
Response: 130262
Conc: 36.78 ng/ml




#44 AR-1268-4

R.T.: 9.359 min
Delta R.T.: -0.004 min
Response: 801437
Conc: 355.86 ng/ml



#44 AR-1268-4

R.T.: 8.323 min
Delta R.T.: -0.007 min
Response: 488019
Conc: 322.56 ng/ml




#45 AR-1268-5

R.T.: 9.691 min
Delta R.T.: -0.005 min
Response: 987885
Conc: 56.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
JCSY-SP2

Manual Integrations
APPROVED

Ankita
6/4/2021 10:50:32 AM



#45 AR-1268-5

R.T.: 8.592 min
Delta R.T.: -0.007 min
Response: 538799
Conc: 51.15 ng/ml