

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0052925\
 Data File : P0111352.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 May 2025 23:45
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 00:18:24 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051525.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 15 07:11:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.681	3.677	359.5E6	321.2E6	61.108	58.924
2) SA Decachlor...	8.721	8.670	274.1E6	96311678	53.099	52.056
Target Compounds						
26) L6 AR-1254-1	5.577	5.555	234.0E6	177.9E6	616.573	548.982
27) L6 AR-1254-2	5.726	5.704	183.5E6	144.0E6	561.440	508.227
28) L6 AR-1254-3	6.131	6.104	300.5E6	231.9E6	571.141	537.477
29) L6 AR-1254-4	6.361	6.332	200.3E6	136.7E6	595.912	552.638
30) L6 AR-1254-5	6.780	6.749	264.3E6	182.0E6	557.014	526.161



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO052925\
Data File : PO111352.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 May 2025 23:45
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 00:18:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051525.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 15 07:11:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.681 min
Delta R.T.: -0.002 min
Response: 359509475
Conc: 61.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.677 min
Delta R.T.: -0.002 min
Response: 321218208
Conc: 58.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.721 min
Delta R.T.: -0.004 min
Response: 274141845
Conc: 53.10 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.670 min
Delta R.T.: -0.004 min
Response: 96311678
Conc: 52.06 ng/ml



#26 AR-1254-1


R.T.: 5.577 min
Delta R.T.: -0.003 min
Response: 234035809
Conc: 616.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.555 min
Delta R.T.: -0.003 min
Response: 177891451
Conc: 548.98 ng/ml




#27 AR-1254-2

R.T.: 5.726 min
Delta R.T.: -0.003 min
Response: 183464842
Conc: 561.44 ng/ml



#27 AR-1254-2


R.T.: 5.704 min
Delta R.T.: -0.002 min
Response: 143952721
Conc: 508.23 ng/ml



#28 AR-1254-3


R.T.: 6.131 min
Delta R.T.: -0.003 min
Response: 300501946
Conc: 571.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.104 min
Delta R.T.: -0.003 min
Response: 231895493
Conc: 537.48 ng/ml




#29 AR-1254-4

R.T.: 6.361 min
Delta R.T.: -0.003 min
Response: 200273135
Conc: 595.91 ng/ml



#29 AR-1254-4


R.T.: 6.332 min
Delta R.T.: -0.003 min
Response: 136691567
Conc: 552.64 ng/ml



#30 AR-1254-5

R.T.: 6.780 min
Delta R.T.: -0.004 min
Response: 264277809
Conc: 557.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.749 min
Delta R.T.: -0.003 min
Response: 182020264
Conc: 526.16 ng/ml