

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0053024\
 Data File : P0104196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2024 17:44
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 18:57:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0053024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu May 30 18:54:17 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.410	3.554	413.5E6	130.5E6	50.000	50.000
2) SA Decachlor...	10.115	8.518	385.8E6	178.9E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.571	4.619	128.4E6	33054672	500.000	500.000
4) L1 AR-1016-2	5.594	4.636	192.4E6	60329200	500.000	500.000
5) L1 AR-1016-3	5.656	4.810	117.9E6	32694966	500.000	500.000
6) L1 AR-1016-4	5.754	4.852	95690363	22845379	500.000	500.000
7) L1 AR-1016-5	6.048	5.063	97949753	32487343	500.000	500.000
31) L7 AR-1260-1	7.171	6.091	194.3E6	63934389	500.000	500.000
32) L7 AR-1260-2	7.427	6.280	226.3E6	77546572	500.000	500.000
33) L7 AR-1260-3	7.786	6.432	178.7E6	74475011	500.000	500.000
34) L7 AR-1260-4	8.011	6.906	205.3E6	63259187	500.000	500.000
35) L7 AR-1260-5	8.326	7.151	393.9E6	164.1E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO053024\
Data File : PO104196.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2024 17:44
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 18:57:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO053024.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu May 30 18:54:17 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.410 min
Delta R.T.: 0.000 min
Response: 413478691
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.554 min
Delta R.T.: 0.000 min
Response: 130523708
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.115 min
Delta R.T.: 0.000 min
Response: 385789748
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.518 min
Delta R.T.: 0.000 min
Response: 178882340
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.571 min
Delta R.T.: 0.000 min
Response: 128409653
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.619 min
Delta R.T.: 0.000 min
Response: 33054672
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.594 min
Delta R.T.: 0.000 min
Response: 192384003
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.636 min
Delta R.T.: 0.000 min
Response: 60329200
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.656 min
Delta R.T.: 0.000 min
Response: 117882482
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.810 min
Delta R.T.: 0.000 min
Response: 32694966
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.754 min
Delta R.T.: 0.000 min
Response: 95690363
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.852 min
Delta R.T.: 0.000 min
Response: 22845379
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.048 min
Delta R.T.: 0.000 min
Response: 97949753
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.063 min
Delta R.T.: 0.000 min
Response: 32487343
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.171 min
Delta R.T.: 0.000 min
Response: 194277285
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.091 min
Delta R.T.: 0.000 min
Response: 63934389
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.427 min
Delta R.T.: 0.000 min
Response: 226299108
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.280 min
Delta R.T.: 0.000 min
Response: 77546572
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.786 min
Delta R.T.: 0.000 min
Response: 178662498
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.432 min
Delta R.T.: 0.000 min
Response: 74475011
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.011 min
Delta R.T.: 0.000 min
Response: 205345786
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 6.906 min
Delta R.T.: 0.000 min
Response: 63259187
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.326 min
Delta R.T.: 0.000 min
Response: 393895634
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.151 min
Delta R.T.: 0.000 min
Response: 164106393
Conc: 500.00 ng/ml