

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0053024\
 Data File : P0104217.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 May 2024 00:51
 Operator : YP/AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 03:59:17 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0053024.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 31 03:58:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.413	3.552	912.5E6	266.0E6	98.936	100.219
2) SA Decachlor...	10.118	8.516	1307.7E6	618.3E6	95.293	95.870
Target Compounds						
41) L9 AR-1268-1	8.636	7.435	1095.8E6	469.4E6	966.762	971.193
42) L9 AR-1268-2	8.729	7.499	1010.5E6	498.5E6	967.274	973.783
43) L9 AR-1268-3	8.956	7.705	890.5E6	450.1E6	967.904	969.879
44) L9 AR-1268-4	9.373	7.993	342.5E6	154.6E6	958.408	958.508
45) L9 AR-1268-5	9.782	8.274	2816.8E6	1379.9E6	978.318	986.279



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO053024\
Data File : PO104217.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 May 2024 00:51
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 03:59:17 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO053024.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 31 03:58:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.413 min
Delta R.T.: 0.000 min
Response: 912549363
Conc: 98.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.552 min
Delta R.T.: 0.000 min
Response: 265969552
Conc: 100.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.118 min
Delta R.T.: 0.000 min
Response: 1307747176
Conc: 95.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.516 min
Delta R.T.: 0.000 min
Response: 618259260
Conc: 95.87 ng/ml



#41 AR-1268-1


R.T.: 8.636 min
Delta R.T.: 0.000 min
Response: 1095795987
Conc: 966.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#41 AR-1268-1

R.T.: 7.435 min
Delta R.T.: 0.000 min
Response: 469403055
Conc: 971.19 ng/ml




#42 AR-1268-2

R.T.: 8.729 min
Delta R.T.: 0.000 min
Response: 1010533798
Conc: 967.27 ng/ml



#42 AR-1268-2


R.T.: 7.499 min
Delta R.T.: 0.000 min
Response: 498527846
Conc: 973.78 ng/ml



#43 AR-1268-3


R.T.: 8.956 min
Delta R.T.: 0.000 min
Response: 890485116
Conc: 967.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#43 AR-1268-3

R.T.: 7.705 min
Delta R.T.: 0.000 min
Response: 450058627
Conc: 969.88 ng/ml




#44 AR-1268-4

R.T.: 9.373 min
Delta R.T.: 0.000 min
Response: 342542843
Conc: 958.41 ng/ml



#44 AR-1268-4


R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 154568916
Conc: 958.51 ng/ml



#45 AR-1268-5

R.T.: 9.782 min
Delta R.T.: 0.000 min
Response: 2816813965
Conc: 978.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 8.274 min
Delta R.T.: 0.000 min
Response: 1379873254
Conc: 986.28 ng/ml