

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060122\
 Data File : P0086821.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Jun 2022 10:46
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 12:46:07 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 01 12:41:09 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.668	4445214	2384075	72.978	71.620
2) SA Decachlor...	10.387	8.702	3482106	1665293	72.862	71.999
Target Compounds						
3) L1 AR-1016-1	5.698	4.762	1486604	708653	702.484	693.385
4) L1 AR-1016-2	5.721	4.781	2162198	1009201	726.730	715.242
5) L1 AR-1016-3	5.785	4.957	1321519	546627	717.144	705.685
6) L1 AR-1016-4	5.884	5.000	1062436	428744	720.086	702.839
7) L1 AR-1016-5	6.181	5.213	1046747	543723	712.998	702.908
31) L7 AR-1260-1	7.317	6.250	1850045	956718	718.520	712.141
32) L7 AR-1260-2	7.575	6.437	2266635	1146115	709.158	711.150
33) L7 AR-1260-3	7.936	6.591	1584848	1150995	721.475	705.684
34) L7 AR-1260-4	8.167	7.065	1872850	859927	718.770	717.462
35) L7 AR-1260-5	8.496	7.305	3582167	2032483	731.189	727.630



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060122\
Data File : P0086821.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Jun 2022 10:46
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 12:46:07 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 01 12:41:09 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: -0.001 min
Response: 4445214
Conc: 72.98 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.668 min
Delta R.T.: 0.000 min
Response: 2384075
Conc: 71.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.387 min
Delta R.T.: 0.002 min
Response: 3482106
Conc: 72.86 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.702 min
Delta R.T.: 0.002 min
Response: 1665293
Conc: 72.00 ng/ml



#3 AR-1016-1


R.T.: 5.698 min
Delta R.T.: 0.000 min
Response: 1486604
Conc: 702.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.762 min
Delta R.T.: 0.000 min
Response: 708653
Conc: 693.38 ng/ml




#4 AR-1016-2

R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 2162198
Conc: 726.73 ng/ml



#4 AR-1016-2


R.T.: 4.781 min
Delta R.T.: 0.000 min
Response: 1009201
Conc: 715.24 ng/ml



#5 AR-1016-3


R.T.: 5.785 min
Delta R.T.: 0.000 min
Response: 1321519
Conc: 717.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.957 min
Delta R.T.: 0.000 min
Response: 546627
Conc: 705.69 ng/ml




#6 AR-1016-4

R.T.: 5.884 min
Delta R.T.: 0.000 min
Response: 1062436
Conc: 720.09 ng/ml



#6 AR-1016-4


R.T.: 5.000 min
Delta R.T.: 0.000 min
Response: 428744
Conc: 702.84 ng/ml



#7 AR-1016-5


R.T.: 6.181 min
Delta R.T.: -0.002 min
Response: 1046747
Conc: 713.00 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.213 min
Delta R.T.: 0.000 min
Response: 543723
Conc: 702.91 ng/ml




#31 AR-1260-1

R.T.: 7.317 min
Delta R.T.: 0.000 min
Response: 1850045
Conc: 718.52 ng/ml



#31 AR-1260-1


R.T.: 6.250 min
Delta R.T.: 0.000 min
Response: 956718
Conc: 712.14 ng/ml



#32 AR-1260-2


R.T.: 7.575 min
Delta R.T.: 0.000 min
Response: 2266635
Conc: 709.16 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 6.437 min
Delta R.T.: 0.000 min
Response: 1146115
Conc: 711.15 ng/ml




#33 AR-1260-3

R.T.: 7.936 min
Delta R.T.: 0.000 min
Response: 1584848
Conc: 721.47 ng/ml



#33 AR-1260-3


R.T.: 6.591 min
Delta R.T.: 0.000 min
Response: 1150995
Conc: 705.68 ng/ml



#34 AR-1260-4


R.T.: 8.167 min
Delta R.T.: 0.001 min
Response: 1872850
Conc: 718.77 ng/ml

Instrument :
ECD_O
ClientSampleId :




#34 AR-1260-4

R.T.: 7.065 min
Delta R.T.: 0.000 min
Response: 859927
Conc: 717.46 ng/ml



#35 AR-1260-5

R.T.: 8.496 min
Delta R.T.: 0.001 min
Response: 3582167
Conc: 731.19 ng/ml



#35 AR-1260-5

R.T.: 7.305 min
Delta R.T.: 0.000 min
Response: 2032483
Conc: 727.63 ng/ml