

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060122\
 Data File : P0086828.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 Jun 2022 12:57
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 14:14:31 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 01 14:13:22 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.668	4635012	2639725	73.360	73.581
2) SA Decachlor...	10.388	8.700	3572211	1760956	72.140	71.435
Target Compounds						
16) L4 AR-1242-1	5.699	4.762	1194087	562735	717.882	696.609
17) L4 AR-1242-2	5.723	4.781	1703595	787130	714.304	702.042
18) L4 AR-1242-3	5.786	4.957	1057179	431628	711.661	712.958
19) L4 AR-1242-4	5.885	5.042	838709	413772	710.566	708.713
20) L4 AR-1242-5	6.627	5.566	889344	500076	718.309	713.273



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060122\
Data File : PO086828.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 Jun 2022 12:57
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 01 14:14:31 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 01 14:13:22 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.000 min
Response: 4635012
Conc: 73.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.668 min
Delta R.T.: 0.000 min
Response: 2639725
Conc: 73.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.388 min
Delta R.T.: 0.002 min
Response: 3572211
Conc: 72.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.700 min
Delta R.T.: 0.000 min
Response: 1760956
Conc: 71.43 ng/ml



#16 AR-1242-1


R.T.: 5.699 min
Delta R.T.: 0.000 min
Response: 1194087
Conc: 717.88 ng/ml

Instrument :
ECD_O
ClientSampleId :




#16 AR-1242-1

R.T.: 4.762 min
Delta R.T.: 0.000 min
Response: 562735
Conc: 696.61 ng/ml




#17 AR-1242-2

R.T.: 5.723 min
Delta R.T.: 0.000 min
Response: 1703595
Conc: 714.30 ng/ml



#17 AR-1242-2


R.T.: 4.781 min
Delta R.T.: 0.000 min
Response: 787130
Conc: 702.04 ng/ml



#18 AR-1242-3


R.T.: 5.786 min
Delta R.T.: 0.000 min
Response: 1057179
Conc: 711.66 ng/ml

Instrument :
ECD_O
ClientSampleId :




#18 AR-1242-3

R.T.: 4.957 min
Delta R.T.: 0.000 min
Response: 431628
Conc: 712.96 ng/ml




#19 AR-1242-4

R.T.: 5.885 min
Delta R.T.: 0.000 min
Response: 838709
Conc: 710.57 ng/ml



#19 AR-1242-4


R.T.: 5.042 min
Delta R.T.: 0.000 min
Response: 413772
Conc: 708.71 ng/ml



#20 AR-1242-5

R.T.: 6.627 min
Delta R.T.: 0.000 min
Response: 889344
Conc: 718.31 ng/ml

Instrument :
ECD_O
ClientSampleId :



#20 AR-1242-5

R.T.: 5.566 min
Delta R.T.: 0.000 min
Response: 500076
Conc: 713.27 ng/ml