

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0060123.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jun 02 06:29:06 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0095419.D 1000 =P0095417.D 750 =P0095418.D
 250 =P0095420.D 50 =P0095421.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.664	3.608	3.668	3.803	3.367	3.622 E6	4.41
2) SA	Decachlorobiphenyl	4.633	4.396	4.548	4.852	4.642	4.614 E6	3.59
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	3.684	3.552	3.655	3.840	3.710	3.688 E5	2.82
42) L9	AR-1268-2	3.363	3.236	3.336	3.506	3.498	3.388 E5	3.38
43) L9	AR-1268-3	2.889	2.776	2.863	3.014	2.913	2.891 E5	2.98
44) L9	AR-1268-4	1.288	1.225	1.269	1.344	1.233	1.272 E5	3.78
45) L9	AR-1268-5	9.448	9.250	9.452	9.642	9.238	9.406 E5	1.78

Signal #2 Calibration Files

500 =P0095419.D 1000 =P0095417.D 750 =P0095418.D
 250 =P0095420.D 50 =P0095421.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.661	1.634	1.663	1.713	1.508	1.636 E6	4.70
2) SA	Decachlorobiphenyl	2.280	2.170	2.241	2.395	2.488	2.315 E6	5.47
3) L1	AR-1016-1						0.000	-1.00
4) L1	AR-1016-2						0.000	-1.00
5) L1	AR-1016-3						0.000	-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0060123.M
 Title : GC EXTRACTABLES
 Last Update : Fri Jun 02 06:29:06 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0095419.D 1000 =P0095417.D 750 =P0095418.D
 250 =P0095420.D 50 =P0095421.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	2.108	2.025	2.102	2.251	2.321	2.161 E5	5.60
42) L9	AR-1268-2	1.887	1.816	1.881	1.979	1.981	1.909 E5	3.72
43) L9	AR-1268-3	1.600	1.535	1.584	1.678	1.708	1.621 E5	4.36
44) L9	AR-1268-4	6.853	6.421	6.721	7.217	7.189	6.880 E4	4.85
45) L9	AR-1268-5	4.795	4.708	4.818	4.893	4.768	4.797 E5	1.42

(#) = Out of Range